

7471B

Linear ▼

 $\mu \text{ Abs.} :$

5534E


A= 0.0000e+000

B= 1.7754e-004

C= 6.5176e-003

Rho= 0.9995003

Accept=Accepted



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	0.003	0.003	-21	0.000	-21					4.11
0.2	0.200	0.194	-0.006	1056	0.0 %	1056					1
2.5	2.500	2.406	-0.094	13515	0.0 %	13515					3
5.0	5.000	5.062	0.062	28477	0.0 %	28477					4
7.5	7.500	7.702	0.202	43345	0.0 %	43345					1
10.0	10.000	9.833	-0.167	55345	0.0 %	55345					3

$$\begin{array}{r} 4.11 \\ - \\ -3 \\ -4 \\ 1 \\ 3 \\ -2 \end{array}$$

LB131292 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 19 Jun 2024 08:39:39

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-21	-	0.0000	7471B	PPB	19 Jun 2024 08:49:09	S	Std
0.2 - 1	50.2	1056	-	0.2000	7471B	PPB	19 Jun 2024 08:51:26	S	Std
2.5 - 1	52.5	13515	-	2.5000	7471B	PPB	19 Jun 2024 08:53:44	S	Std
5.0 - 1	55	28477	-	5.0000	7471B	PPB	19 Jun 2024 08:56:00	S	Std
7.5 - 1	57.5	43345	-	7.5000	7471B	PPB	19 Jun 2024 08:58:19	S	Std
10.0 - 1	510	55345	-	10.0000	7471B	PPB	19 Jun 2024 09:00:38	S	Std
ICV94 - 1	ICV94	21951	3.9038	-	7471B	PPB	19 Jun 2024 09:07:32	U	SMPL
ICB94 - 1	ICB94	-178	-0.0251	-	7471B	PPB	19 Jun 2024 09:09:47	U	SMPL
CCV54 - 1	CCV54	29069	5.1675	-	7471B	PPB	19 Jun 2024 09:12:07	U	SMPL
CCB54 - 1	CCB54	-362	-0.0578	-	7471B	PPB	19 Jun 2024 09:14:21	U	SMPL
CRA - 1	CRA	959	0.1768	-	7471B	PPB	19 Jun 2024 09:16:40	U	SMPL
HighStd - 1	HighStd	55345	9.8326	-	7471B	PPB	19 Jun 2024 09:18:55	U	SMPL
ChkStd - 1	ChkStd	41078	7.2996	-	7471B	PPB	19 Jun 2024 09:21:11	U	SMPL
PB161540BL - 1	PBS	-371	-0.0594	-	7471B	PPB	19 Jun 2024 09:23:36	U	SMPL
PB161540BS - 1	LCSS	20696	3.6810	-	7471B	PPB	19 Jun 2024 09:25:57	U	SMPL
P2880-01 - 1	OR-03-06132024	21681	3.8558	-	7471B	PPB	19 Jun 2024 09:28:13	U	SMPL
P2880-01DUP - 1	OR-03-06132024DUP	18055	3.2121	-	7471B	PPB	19 Jun 2024 09:30:34	U	SMPL
P2880-01MSD - 1	OR-03-06132024MSD	41249	7.3300	-	7471B	PPB	19 Jun 2024 09:35:14	U	SMPL
P2883-01 - 1	Soil/Gravel-SP	1459	0.2656	-	7471B	PPB	19 Jun 2024 09:37:36	U	SMPL
CCV55 - 1	CCV55	27972	4.9728	-	7471B	PPB	19 Jun 2024 09:39:57	U	SMPL
CCB55 - 1	CCB55	-245	-0.0370	-	7471B	PPB	19 Jun 2024 09:42:14	U	SMPL
P2885-01 - 1	CL-02-061424	4718	0.8442	-	7471B	PPB	19 Jun 2024 09:44:34	U	SMPL
P2887-01 - 1	HD-01-06132024	575	0.1086	-	7471B	PPB	19 Jun 2024 09:46:49	U	SMPL
P2915-01 - 1	WC-NORTH-1	733	0.1367	-	7471B	PPB	19 Jun 2024 09:49:06	U	SMPL
P2915-02 - 1	WC-SOUTH-1	728	0.1358	-	7471B	PPB	19 Jun 2024 09:51:22	U	SMPL
P2916-01 - 1	WCS-TPF	1310	0.2391	-	7471B	PPB	19 Jun 2024 09:53:38	U	SMPL
P2917-01 - 1	ROAD	4060	0.7273	-	7471B	PPB	19 Jun 2024 09:55:54	U	SMPL
P2918-01 - 1	POLE	4554	0.8150	-	7471B	PPB	19 Jun 2024 09:58:11	U	SMPL
P2880-01LX5 - 1		4105	0.7353	-	7471B	PPB	19 Jun 2024 10:00:30	U	SMPL
P2880-01A - 1		43391	7.7103	-	7471B	PPB	19 Jun 2024 10:05:51	U	SMPL
P2880-01MS - 1	OR-03-06132024MS	41559	7.3850	-	7471B	PPB	19 Jun 2024 10:09:00	U	SMPL
CCV56 - 1	CCV56	28782	5.1166	-	7471B	PPB	19 Jun 2024 10:11:55	U	SMPL
CCB56 - 1	CCB56	-275	-0.0423	-	7471B	PPB	19 Jun 2024 10:14:10	U	SMPL