

LB131297


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

40075



A= 0.0000e+000

B= 2.4412e-004

C= 3.0321e-003

Rho= 0.9990061

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	σ/μ
0.0	0.000	0.017	0.017	58	0.000	58	0				-
0.2	0.200	0.196	-0.004	789	0.0 %	789					-2
2.5	2.500	2.440	-0.060	9982	0.0 %	9982					-2
5.0	5.000	4.933	-0.067	20193	0.0 %	20193					-1
7.5	7.500	7.829	0.329	32057	0.0 %	32057					4
10.0	10.000	9.786	-0.214	40075	0.0 %	40075					-2

LB131297 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 19 Jun 2024 11:31:22

Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc	Method	Units	Date	Type	Type
0.0 - 1	50	58	-	0.0000	7470A	PPB	19 Jun 2024 11:44:18	S	Std
0.2 - 1	50.2	789	-	0.2000	7470A	PPB	19 Jun 2024 11:46:35	S	Std
2.5 - 1	52.5	9982	-	2.5000	7470A	PPB	19 Jun 2024 11:48:52	S	Std
5.0 - 1	55	20193	-	5.0000	7470A	PPB	19 Jun 2024 11:51:09	S	Std
7.5 - 1	57.5	32057	-	7.5000	7470A	PPB	19 Jun 2024 11:53:27	S	Std
10.0 - 1	514	40075	-	10.0000	7470A	PPB	19 Jun 2024 11:58:30	S	Std
ICV95 - 1	ICV95	17698	4.3234	-	7470A	PPB	19 Jun 2024 12:08:31	U	SMPL
ICB95 - 1	ICB95	-195	-0.0446	-	7470A	PPB	19 Jun 2024 12:10:46	U	SMPL
CCV57 - 1	CCV57	19688	4.8092	-	7470A	PPB	19 Jun 2024 12:13:02	U	SMPL
CCB57 - 1	CCB57	-249	-0.0578	-	7470A	PPB	19 Jun 2024 12:15:17	U	SMPL
CRA - 1	CRA	719	0.1786	-	7470A	PPB	19 Jun 2024 12:17:35	U	SMPL
HighStd - 1	HighStd	40644	9.9250	-	7470A	PPB	19 Jun 2024 12:19:50	U	SMPL
ChkStd - 1	ChkStd	28492	6.9585	-	7470A	PPB	19 Jun 2024 12:22:05	U	SMPL
PB161541BL - 1	PBW	-389	-0.0919	-	7470A	PPB	19 Jun 2024 12:24:25	U	SMPL
PB161541BS - 1	LCSW	16611	4.0581	-	7470A	PPB	19 Jun 2024 12:26:43	U	SMPL
P2883-03 - 1	Soil/Gravel-SP	-232	-0.0536	-	7470A	PPB	19 Jun 2024 12:29:00	U	SMPL
P2915-03 - 1	WC-NORTH-1	-23	-0.0026	-	7470A	PPB	19 Jun 2024 12:31:18	U	SMPL
P2915-03DUP - 1	WC-NORTH-1DUP	-11	0.0003	-	7470A	PPB	19 Jun 2024 12:33:36	U	SMPL
P2915-03MS - 1	WC-NORTH-1MS	14116	3.4490	-	7470A	PPB	19 Jun 2024 12:35:53	U	SMPL
P2915-03MSD - 1	WC-NORTH-1MSD	13278	3.2444	-	7470A	PPB	19 Jun 2024 12:38:09	U	SMPL
CCV58 - 1	CCV58	20319	4.9633	-	7470A	PPB	19 Jun 2024 12:40:26	U	SMPL
CCB58 - 1	CCB58	-276	-0.0643	-	7470A	PPB	19 Jun 2024 12:42:44	U	SMPL
P2915-04 - 1	WC-SOUTH-1	-18	-0.0014	-	7470A	PPB	19 Jun 2024 12:45:02	U	SMPL
P2916-04 - 1	WCS-TPF	-20	-0.0019	-	7470A	PPB	19 Jun 2024 12:47:17	U	SMPL
P2917-03 - 1	ROAD	-14	-0.0004	-	7470A	PPB	19 Jun 2024 12:49:33	U	SMPL
P2918-03 - 1	POLE	-8	0.0011	-	7470A	PPB	19 Jun 2024 12:51:49	U	SMPL
PB161530TB - 1	PB161530TB	11	0.0057	-	7470A	PPB	19 Jun 2024 12:54:05	U	SMPL
P2915-03LX5 - 1		41	0.0130	-	7470A	PPB	19 Jun 2024 12:56:21	U	SMPL
P2915-03A - 1		14603	3.5679	-	7470A	PPB	19 Jun 2024 12:58:37	U	SMPL
CCV59 - 1	CCV59	20606	5.0333	-	7470A	PPB	19 Jun 2024 13:00:54	U	SMPL
CCB59 - 1	CCB59	-260	-0.0604	-	7470A	PPB	19 Jun 2024 13:03:12	U	SMPL