

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB131298

Ambient Barometric Pressure (mmHg): 776.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3085

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.4	06/19/2024	08:20
2	P2938-04		1	12	100.00	>212.0	>212.0	06/19/2024	08:45
3	P2938-04DUP		1	12	100.00	>212.0	>212.0	06/19/2024	09:15
4	P2938-05		1	14	100.00	>212.0	>212.0	06/19/2024	10:30
5	P2938-06		1	11	100.00	>212.0	>212.0	06/19/2024	11:00
6	P2938-07		1	12	100.00	>212.0	>212.0	06/19/2024	12:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB131298

WorkList Name : FP-6-19

WorkList ID : 181197

Department : Wet-Chemistry

Date : 06-19-2024 07:52:32

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2938-04	399-400	Water	Flash Point	Cool 4 deg C	PSEG03	M21	06/18/2024	1010B
P2938-05	401-402	Water	Flash Point	Cool 4 deg C	PSEG03	M21	06/18/2024	1010B
P2938-06	403	Water	Flash Point	Cool 4 deg C	PSEG03	M21	06/18/2024	1010B
P2938-07	405	Water	Flash Point	Cool 4 deg C	PSEG03	M21	06/18/2024	1010B

Date/Time 06/19/2024 08:00
Raw Sample Received by: RM
Raw Sample Relinquished by: RM

Date/Time 06/19/2024 12:45
Raw Sample Received by: RM
Raw Sample Relinquished by: RM