

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB131303

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	06/19/2024	11:40
2	CAL2	1	Water	NA	NA	20.2	7.00	06/19/2024	11:41
3	CAL3	1	Water	NA	NA	20.3	10.02	06/19/2024	11:42
4	ICV	1	Water	NA	NA	20.4	7.02	06/19/2024	11:45
5	CCV1	1	Water	NA	NA	20.3	2.01	06/19/2024	11:47
6	P2956-02	1	Water	NA	NA	20.2	7.60	06/19/2024	11:55
7	P2956-02DUP	1	Water	NA	NA	20.3	7.61	06/19/2024	11:56
8	CCV2	1	Water	NA	NA	20.3	12.02	06/19/2024	12:00

WORKLIST(Hardcopy Internal Chain)

131303

WorkList Name : corrosivity p2956 water WorkList ID : 181206 Department : Wet-Chemistry Date : 06-19-2024 11:12:51

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2956-02	WC-W-01-202406	Water	Corrosivity	Cool 4 deg C	AECO02	M21	06/18/2024	9040C

Date/Time 06.19.24 11:30
Raw Sample Received by: JH GRC
Raw Sample Relinquished by: JH GRC

Date/Time 06.19.24 13:00
Raw Sample Received by: JH GRC
Raw Sample Relinquished by: JH GRC