

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 6/24/2024

OVENTEMP IN Celsius(°C): 108
Time IN: 14:40
In Date: 06/19/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:11
Out Date: 06/20/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB131306

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
P2872-01	MA4C66	1	1.17	8.82	9.99	8.22	79.9	
P2872-02	MA4C66D	2	1.17	8.82	9.99	8.22	79.9	
P2872-03	MA4C66S	3	1.17	8.82	9.99	8.22	79.9	
P2872-04	MA4C51	4	1.17	8.40	9.57	7.68	77.5	
P2872-05	MA4CN6	5	1.15	8.39	9.54	8.04	82.1	
P2872-06	MA4CN8	6	1.15	8.73	9.88	9.6	96.8	
P2872-07	MA4B58	7	1.14	8.74	9.88	7.67	74.7	
P2872-08	MA4B59	8	1.15	8.83	9.98	8.55	83.8	
P2872-09	MA4B73	9	1.15	8.84	9.99	4.75	40.7	
P2872-10	MA4CP8	10	1.15	8.79	9.94	5.49	49.4	
P2872-11	MA4CP9	11	1.18	8.79	9.97	6.96	65.8	
P2872-12	MA4CQ0	12	1.13	8.63	9.76	6.66	64.1	
P2872-13	MA4CQ1	13	1.12	8.50	9.62	6.72	65.9	
P2872-14	MA4CQ2	14	1.16	8.50	9.66	7.2	71.1	
P2872-15	MA4B56	15	1.18	8.66	9.84	6.68	63.5	
P2872-16	MA4CP0	16	1.15	8.80	9.95	6.64	62.4	
P2872-17	MA4CP1	17	1.14	8.84	9.98	6.81	64.1	
P2872-18	MA4CP2	18	1.12	8.71	9.83	7.78	76.5	
P2872-19	MA4CP3	19	1.14	8.80	9.94	8.77	86.7	
P2872-20	MA4CQ3	20	1.12	8.77	9.89	5.5	49.9	
P2872-21	MA4CQ4	21	1.15	8.82	9.97	5.93	54.2	
P2872-22	MA4CQ5	22	1.12	8.86	9.98	8.3	81.0	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$