

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB131315

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	06/19/2024	14:00
2	CAL2	1	Water	NA	NA	20.3	7.00	06/19/2024	14:02
3	CAL3	1	Water	NA	NA	20.3	10.02	06/19/2024	14:05
4	ICV	1	Water	NA	NA	20.3	7.02	06/19/2024	14:10
5	CCV1	1	Water	NA	NA	20.2	2.01	06/19/2024	14:11
6	P2976-04	1	Solid	20.02	20	23.9	5.50	06/19/2024	14:15
7	P2976-04DUP	1	Solid	20.03	20	23.8	5.51	06/19/2024	14:16
8	CCV2	1	Water	NA	NA	20.3	12.02	06/19/2024	14:20

WORKLIST(Hardcopy Internal Chain)

WB 131313

WorkList Name : corrosivity p2976 WorkList ID : 181213 Department : Wet-Chemistry Date : 06-19-2024 13:48:40

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2976-04	MH-K-COMP	Solid	Corrosivity	Cool 4 deg C	PSEG03	M21	06/19/2024	9045D

Date/Time 06.19.24 13150
Raw Sample Received by: [Signature] 69C
Raw Sample Relinquished by: [Signature]

Date/Time 06.19.24 16130
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]