

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB131318

Ambient Barometric Pressure (mmHg): 776.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3085

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.4	06/20/2024	08:00
2	P2879-01		1	12	100.00	>212.0	>212.0	06/20/2024	08:30
3	P2879-01DUP		1	12	100.00	>212.0	>212.0	06/20/2024	09:00
4	P2942-01		1	14	100.00	>212.0	>212.0	06/20/2024	09:30
5	P2956-02		1	11	100.00	>212.0	>212.0	06/20/2024	10:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6/13/18

WorkList Name : fp6-19

WorkList ID : 181221

Department : Wet-Chemistry

Date : 06-19-2024 16:46:58

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2879-01	50363	Water	Flash Point	Cool 4 deg C	PSEG03	M31	06/14/2024	1010B
P2942-01	GRAY-WATER	Water	Flash Point	Cool 4 deg C	PSEG03	M21	06/18/2024	1010B
P2956-02	WC-W-01-202406	Water	Flash Point	Cool 4 deg C	AECO02	M21	06/18/2024	1010B

Date/Time 26/20/2024 07:50
 Raw Sample Received by: RM CWCJ
 Raw Sample Relinquished by: M CWCJ

Date/Time 06/20/2024 10:40
 Raw Sample Received by: M CWCJ
 Raw Sample Relinquished by: RM CWCJ