

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 6/21/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 12:35
In Date: 06/20/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:36
Out Date: 06/21/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB131325

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
P2871-01	A4C66	1	1.17	8.82	9.99	8.22	79.9	
P2871-02	A4C66MS	2	1.17	8.82	9.99	8.22	79.9	
P2871-03	A4C66MSD	3	1.17	8.82	9.99	8.22	79.9	
P2871-04	A4C52	4	1.18	8.78	9.96	9.16	90.9	
P2871-05	A4C53	5	1.15	8.50	9.65	8.95	91.8	
P2871-06	A4C54	6	1.12	8.77	9.89	6.71	63.7	
P2871-07	A4C55	7	1.16	8.82	9.98	7.66	73.7	
P2871-08	A4C57	8	1.18	8.42	9.6	8.78	90.3	
P2871-09	A4C58	9	1.15	8.75	9.9	9.09	90.7	
P2871-10	A4C60	10	1.19	8.79	9.98	8.66	85.0	
P2871-11	A4CJ1	11	1.16	8.69	9.85	5.08	45.1	
P2871-12	A4CJ2	12	1.19	8.50	9.69	8.27	83.3	
P2871-13	A4CN3	13	1.16	8.82	9.98	9.22	91.4	
P2871-14	A4CN4	14	1.18	8.33	9.51	6.45	63.3	
P2871-15	A4CN5	15	1.19	8.53	9.72	7.36	72.3	
P2871-16	A4CN7	16	1.17	8.55	9.72	8.92	90.6	
P2871-17	A4CN8	17	1.14	8.74	9.88	9.6	96.8	
P2871-18	A4DE4	18	1.00	1.00	2.00	2.00	100.0	E.B.
P2871-19	VHBLK001	19	1.00	1.00	2.00	2.00	100.0	vhblk
P2871-20	A4CN6	20	1.15	8.39	9.54	8.04	82.1	
P2871-21	A4DN7	21	1.00	1.00	2.00	2.00	100.0	P.T. SAMPLE
P2871-22	A4B53	22	1.13	8.55	9.68	5.97	56.6	
P2871-23	A4B54	23	1.12	8.70	9.82	7.76	76.3	
P2871-24	A4DD7	24	1.00	1.00	2.00	2.00	100.0	E.B.

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$