

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 6/21/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 14:20
In Date: 06/20/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:40
Out Date: 06/21/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB131329

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
P2947-01	EZXD2	1	1.18	8.72	9.9	8.28	81.4	
P2947-02	EZXD3	2	1.18	8.68	9.86	8.64	85.9	
P2947-03	EZXD4	3	1.18	8.54	9.72	8.29	83.3	
P2947-04	EZXD5	4	1.14	8.71	9.85	8.35	82.8	
P2947-05	EZXD6	5	1.15	8.80	9.95	8.42	82.6	
P2947-06	EZXD7	6	1.18	8.55	9.73	8.13	81.3	
P2947-07	EZXD7D	7	1.18	8.55	9.73	8.13	81.3	
P2947-08	EZXD7S	8	1.18	8.55	9.73	8.13	81.3	
P2947-09	EZXD8	9	1.14	8.63	9.77	8.36	83.7	
P2947-10	EZXD9	10	1.12	8.60	9.72	8.26	83.0	
P2947-11	EZXEO	11	1.15	8.51	9.66	8.07	81.3	
P2947-12	EZXEO1	12	1.18	8.73	9.91	8.2	80.4	
P2947-13	EZXEO2	13	1.15	8.40	9.55	8.83	91.4	
P2947-14	EZXEO3	14	1.15	8.80	9.95	8.59	84.5	
P2947-15	EZXEO4	15	1.17	8.58	9.75	8.52	85.7	
P2947-16	EZXEO5	16	1.15	8.84	9.99	8.49	83.0	
P2947-17	EZXEO6	17	1.18	8.50	9.68	8.61	87.4	
P2947-18	EZXEO7	18	1.19	8.62	9.81	8.17	81.0	
P2947-19	EZXEO8	19	1.14	8.85	9.99	8.37	81.7	
P2947-20	E10EO	20	1.18	8.81	9.99	8.39	81.8	
P2947-21	E10EO1	21	1.18	8.74	9.92	8.18	80.1	
P2947-22	E10EO2	22	1.15	8.38	9.53	8.25	84.7	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$