

**PERCENT SOLID**

**Supervisor:** Iwona  
**Analyst:** jignesh  
**Date:** 6/21/2024

**OVENTEMP IN Celsius(°C):** 107  
**Time IN:** 17:00  
**In Date:** 06/20/2024  
**Weight Check 1.0g:** 1.00  
**Weight Check 10g:** 10.00  
**OvenID:** M OVEN#1

**OVENTEMP OUT Celsius(°C):** 103  
**Time OUT:** 08:00  
**Out Date:** 06/21/2024  
**Weight Check 1.0g:** 1.00  
**Weight Check 10g:** 10.00  
**BalanceID:** M SC-4  
**Thermometer ID:** % SOLID- OVEN

QC:LB131333

| Lab ID   | Client SampleID | Dish # | Dish Wt (g) (A) | Sample Wt (g) | Dish + Sample Wt (g) (B) | Dish+Dry Sample Wt (g) (C) | % Solid | Comments       |
|----------|-----------------|--------|-----------------|---------------|--------------------------|----------------------------|---------|----------------|
| P2926-03 | OILY-GRAVEL     | 1      | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | OILY-GRAVEL    |
| P2977-01 | OSTL-A1         | 2      | 1.18            | 8.33          | 9.51                     | 8.00                       | 81.9    |                |
| P2977-02 | OSTL-A2         | 3      | 1.18            | 8.32          | 9.5                      | 8.29                       | 85.5    |                |
| P2977-03 | OSTL-A3         | 4      | 1.18            | 8.44          | 9.62                     | 8.22                       | 83.4    |                |
| P2977-04 | OSTL-A4         | 5      | 1.16            | 8.68          | 9.84                     | 8.23                       | 81.5    |                |
| P2977-05 | OSTL-A5         | 6      | 1.18            | 8.56          | 9.74                     | 8.48                       | 85.3    |                |
| P2977-06 | OSTL-A6         | 7      | 1.16            | 8.65          | 9.81                     | 8.43                       | 84.0    |                |
| P2977-07 | OSTL-B1         | 8      | 1.18            | 8.52          | 9.7                      | 8.17                       | 82.0    |                |
| P2977-08 | OSTL-B2         | 9      | 1.12            | 8.76          | 9.88                     | 8.36                       | 82.6    |                |
| P2977-09 | OSTL-B3         | 10     | 1.18            | 8.67          | 9.85                     | 8.21                       | 81.1    |                |
| P2977-10 | OSTL-B4         | 11     | 1.16            | 8.71          | 9.87                     | 8.51                       | 84.4    |                |
| P2977-11 | OSTL-B5         | 12     | 1.15            | 8.82          | 9.97                     | 8.16                       | 79.5    |                |
| P2977-12 | OSTL-B6         | 13     | 1.15            | 8.40          | 9.55                     | 8.02                       | 81.8    |                |
| P2977-13 | OSTL-C1         | 14     | 1.19            | 8.70          | 9.89                     | 8.38                       | 82.6    |                |
| P2977-14 | OSTL-C2         | 15     | 1.13            | 8.70          | 9.83                     | 8.31                       | 82.5    |                |
| P2977-15 | OSTL-C3         | 16     | 1.14            | 8.80          | 9.94                     | 8.3                        | 81.4    |                |
| P2977-16 | OSTL-C4         | 17     | 1.18            | 8.56          | 9.74                     | 9.00                       | 91.4    |                |
| P2977-17 | OSTL-C5         | 18     | 1.16            | 8.48          | 9.64                     | 7.97                       | 80.3    |                |
| P2977-18 | OSTL-C6         | 19     | 1.15            | 8.82          | 9.97                     | 8.4                        | 82.2    |                |
| P2977-19 | OSTL-D1         | 20     | 1.15            | 8.40          | 9.55                     | 7.85                       | 79.8    |                |
| P2977-20 | OSTL-D2         | 21     | 1.18            | 8.57          | 9.75                     | 7.93                       | 78.8    |                |
| P2977-21 | OSTL-D3         | 22     | 1.15            | 8.82          | 9.97                     | 8.8                        | 86.7    |                |
| P2977-22 | OSTL-D4         | 23     | 1.18            | 8.34          | 9.52                     | 8.31                       | 85.5    |                |
| P2977-23 | OSTL-D5         | 24     | 1.19            | 8.77          | 9.96                     | 8.67                       | 85.3    |                |
| P2977-24 | OSTL-D6         | 25     | 1.18            | 8.45          | 9.63                     | 8.41                       | 85.6    |                |
| P2984-01 | PCB061424RB01   | 26     | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | WINDOW SEALANT |
| P2994-03 | 32518           | 34     | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | OIL SAMPLE     |
| P2995-01 | HR-02-06202024  | 27     | 1.18            | 8.43          | 9.61                     | 9.27                       | 96.0    |                |

**PERCENT SOLID**

**Supervisor:** Iwona  
**Analyst:** jignesh  
**Date:** 6/21/2024

**OVENTEMP IN Celsius(°C):** 107  
**Time IN:** 17:00  
**In Date:** 06/20/2024  
**Weight Check 1.0g:** 1.00  
**Weight Check 10g:** 10.00  
**OvenID:** M OVEN#1

**OVENTEMP OUT Celsius(°C):** 103  
**Time OUT:** 08:00  
**Out Date:** 06/21/2024  
**Weight Check 1.0g:** 1.00  
**Weight Check 10g:** 10.00  
**BalanceID:** M SC-4  
**Thermometer ID:** % SOLID- OVEN

QC:LB131333

| Lab ID   | Client SampleID   | Dish # | Dish Wt (g) (A) | Sample Wt (g) | Dish + Sample Wt (g) (B) | Dish+Dry Sample Wt (g) (C) | % Solid | Comments |
|----------|-------------------|--------|-----------------|---------------|--------------------------|----------------------------|---------|----------|
| P2995-02 | HR-02-06202024-E2 | 28     | 1.18            | 8.54          | 9.72                     | 9.03                       | 91.9    |          |
| P2997-01 | TR-06-062024      | 29     | 1.13            | 8.77          | 9.9                      | 9.14                       | 91.3    |          |
| P2997-02 | TR-06-062024-E2   | 30     | 1.13            | 8.60          | 9.73                     | 9.06                       | 92.2    |          |
| P2998-01 | WCS-TPL           | 31     | 1.18            | 8.44          | 9.62                     | 9.07                       | 93.5    |          |
| P2998-02 | WCS-TPL-EPH       | 32     | 1.16            | 8.53          | 9.69                     | 9.14                       | 93.6    |          |
| P2998-03 | WCS-TPL-VOC       | 33     | 1.19            | 8.52          | 9.71                     | 9.16                       | 93.5    |          |

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

# WORKLIST(Hardcopy Internal Chain)

131333

WorkList Name : %1-062024      WorkList ID : 181226      Department : Wet-Chemistry      Date : 06-20-2024 08:08:02

| Sample   | Customer Sample | Matrix | Test           | Preservative | Customer | Raw Sample Storage Location | Collect Date | Method       |
|----------|-----------------|--------|----------------|--------------|----------|-----------------------------|--------------|--------------|
| P2926-03 | OILY-GRAVEL     | Solid  | Percent Solids | Cool 4 deg C | PSEG03   | M21                         | 06/17/2024   | Chemtech -SO |
| P2977-01 | OSTL-A1         | Solid  | Percent Solids | Cool 4 deg C | JPCL01   | M31                         | 06/19/2024   | Chemtech -SO |
| P2977-02 | OSTL-A2         | Solid  | Percent Solids | Cool 4 deg C | JPCL01   | M31                         | 06/19/2024   | Chemtech -SO |
| P2977-03 | OSTL-A3         | Solid  | Percent Solids | Cool 4 deg C | JPCL01   | M31                         | 06/19/2024   | Chemtech -SO |
| P2977-04 | OSTL-A4         | Solid  | Percent Solids | Cool 4 deg C | JPCL01   | M31                         | 06/19/2024   | Chemtech -SO |
| P2977-05 | OSTL-A5         | Solid  | Percent Solids | Cool 4 deg C | JPCL01   | M31                         | 06/19/2024   | Chemtech -SO |
| P2977-06 | OSTL-A6         | Solid  | Percent Solids | Cool 4 deg C | JPCL01   | M31                         | 06/19/2024   | Chemtech -SO |
| P2977-07 | OSTL-B1         | Solid  | Percent Solids | Cool 4 deg C | JPCL01   | M31                         | 06/19/2024   | Chemtech -SO |
| P2977-08 | OSTL-B2         | Solid  | Percent Solids | Cool 4 deg C | JPCL01   | M31                         | 06/19/2024   | Chemtech -SO |
| P2977-09 | OSTL-B3         | Solid  | Percent Solids | Cool 4 deg C | JPCL01   | M31                         | 06/19/2024   | Chemtech -SO |
| P2977-10 | OSTL-B4         | Solid  | Percent Solids | Cool 4 deg C | JPCL01   | M31                         | 06/19/2024   | Chemtech -SO |
| P2977-11 | OSTL-B5         | Solid  | Percent Solids | Cool 4 deg C | JPCL01   | M31                         | 06/19/2024   | Chemtech -SO |
| P2977-12 | OSTL-B6         | Solid  | Percent Solids | Cool 4 deg C | JPCL01   | M31                         | 06/19/2024   | Chemtech -SO |
| P2977-13 | OSTL-C1         | Solid  | Percent Solids | Cool 4 deg C | JPCL01   | M31                         | 06/19/2024   | Chemtech -SO |
| P2977-14 | OSTL-C2         | Solid  | Percent Solids | Cool 4 deg C | JPCL01   | M31                         | 06/19/2024   | Chemtech -SO |
| P2977-15 | OSTL-C3         | Solid  | Percent Solids | Cool 4 deg C | JPCL01   | M31                         | 06/19/2024   | Chemtech -SO |
| P2977-16 | OSTL-C4         | Solid  | Percent Solids | Cool 4 deg C | JPCL01   | M31                         | 06/19/2024   | Chemtech -SO |
| P2977-17 | OSTL-C5         | Solid  | Percent Solids | Cool 4 deg C | JPCL01   | M31                         | 06/19/2024   | Chemtech -SO |
| P2977-18 | OSTL-C6         | Solid  | Percent Solids | Cool 4 deg C | JPCL01   | M31                         | 06/19/2024   | Chemtech -SO |
| P2977-19 | OSTL-D1         | Solid  | Percent Solids | Cool 4 deg C | JPCL01   | M31                         | 06/19/2024   | Chemtech -SO |
| P2977-20 | OSTL-D2         | Solid  | Percent Solids | Cool 4 deg C | JPCL01   | M31                         | 06/19/2024   | Chemtech -SO |

Date/Time 06/20/24 15:25      Date/Time 06/20/24      17110  
 Raw Sample Received by: [Signature]      Raw Sample Received by: [Signature]  
 Raw Sample Relinquished by: [Signature]      Raw Sample Relinquished by: [Signature]

# WORKLIST(Hardcopy Internal Chain)

131333

WorkList Name : %1-062024      WorkList ID : 181226      Department : Wet-Chemistry      Date : 06-20-2024 08:08:02

| Sample   | Customer Sample   | Matrix | Test           | Preservative | Customer | Raw Sample Storage Location | Collect Date | Method       |
|----------|-------------------|--------|----------------|--------------|----------|-----------------------------|--------------|--------------|
| P2977-21 | OSTL-D3           | Solid  | Percent Solids | Cool 4 deg C | JPCL01   | M31                         | 06/19/2024   | Chemtech -SO |
| P2977-22 | OSTL-D4           | Solid  | Percent Solids | Cool 4 deg C | JPCL01   | M31                         | 06/19/2024   | Chemtech -SO |
| P2977-23 | OSTL-D5           | Solid  | Percent Solids | Cool 4 deg C | JPCL01   | M31                         | 06/19/2024   | Chemtech -SO |
| P2977-24 | OSTL-D6           | Solid  | Percent Solids | Cool 4 deg C | JPCL01   | M31                         | 06/19/2024   | Chemtech -SO |
| P2984-01 | PCB061424RB01     | Solid  | Percent Solids | Cool 4 deg C | CLOU03   | M21                         | 06/14/2024   | Chemtech -SO |
| P2994-03 | 32518             | Solid  | Percent Solids | Cool 4 deg C | PSEG03   | M21                         | 06/20/2024   | Chemtech -SO |
| P2995-01 | HR-02-06202024    | Solid  | Percent Solids | Cool 4 deg C | PSEG05   | M21                         | 06/20/2024   | Chemtech -SO |
| P2995-02 | HR-02-06202024-E2 | Solid  | Percent Solids | Cool 4 deg C | PSEG05   | M21                         | 06/20/2024   | Chemtech -SO |
| P2997-01 | TR-06-062024      | Solid  | Percent Solids | Cool 4 deg C | PSEG05   | M21                         | 06/20/2024   | Chemtech -SO |
| P2997-02 | TR-06-062024-E2   | Solid  | Percent Solids | Cool 4 deg C | PSEG05   | M21                         | 06/20/2024   | Chemtech -SO |
| P2998-01 | WCS-TPL           | Solid  | Percent Solids | Cool 4 deg C | PSEG03   | M21                         | 06/20/2024   | Chemtech -SO |
| P2998-02 | WCS-TPL-EPH       | Solid  | Percent Solids | Cool 4 deg C | PSEG03   | M21                         | 06/20/2024   | Chemtech -SO |
| P2998-03 | WCS-TPL-VOC       | Solid  | Percent Solids | Cool 4 deg C | PSEG03   | M21                         | 06/20/2024   | Chemtech -SO |

Date/Time 06/20/24 15:25  
 Raw Sample Received by: [Signature]  
 Raw Sample Relinquished by: [Signature]

Date/Time 06/20/24 17:10  
 Raw Sample Received by: [Signature]  
 Raw Sample Relinquished by: [Signature]