

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB131335

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	06/20/2024	15:00
2	CAL2	1	Water	NA	NA	20.2	7.00	06/20/2024	15:01
3	CAL3	1	Water	NA	NA	20.3	10.02	06/20/2024	15:05
4	ICV	1	Water	NA	NA	20.3	7.01	06/20/2024	15:10
5	CCV1	1	Water	NA	NA	20.3	2.01	06/20/2024	15:15
6	P2993-05	1	Water	NA	NA	24.7	6.30	06/20/2024	15:25
7	P2993-05DUP	1	Water	NA	NA	24.8	6.31	06/20/2024	15:26
8	P2996-01	1	Water	NA	NA	21.9	6.58	06/20/2024	15:30
9	P2996-02	1	Water	NA	NA	21.6	6.75	06/20/2024	15:37
10	P2996-03	1	Water	NA	NA	22.1	6.92	06/20/2024	15:42
11	CCV2	1	Water	NA	NA	20.3	12.02	06/20/2024	15:47

WB 131335

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph p2996 WorkList ID : 181252 Department : Wet-Chemistry Date : 06-20-2024 14:42:19

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2993-05	453	Water	pH	Cool 4 deg C	PSEG03	M21	06/20/2024	9040C
P2996-01	33710	Water	pH	Cool 4 deg C	PSEG03	M21	06/20/2024	9040C
P2996-02	33711	Water	pH	Cool 4 deg C	PSEG03	M21	06/20/2024	9040C
P2996-03	34342	Water	pH	Cool 4 deg C	PSEG03	M21	06/20/2024	9040C

14:50

Date/Time 06/20/24
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 06/20/24
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]