

LB131349


7471B

INSTRUMENT 101 (V1

Linear

 μ Abs.:

61093



A= 0.0000e+000

B= 1.6305e-004

C= 2.1305e-002

Rho= 0.9999686

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/11)
0.0	0.000	0.030	0.030	51	0.000	51					
0.2	0.200	0.181	-0.019	982	0.0 %	982					
2.5	2.500	2.456	-0.044	14930	0.0 %	14930					
5.0	5.000	5.039	0.039	30772	0.0 %	30772					
7.5	7.500	7.512	0.012	45941	0.0 %	45941					
10.0	10.000	9.983	-0.017	61093	0.0 %	61093					

0/11)

2

16

12

1

0

0

LB131349 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 21 Jun 2024 11:36:55

Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc	Method	Units	Date	Type	Type
0.0 - 1	SO	51	-	0.0000	7471B	PPB	21 Jun 2024 11:42:55	S	Std
0.2 - 1	SO.2	982	-	0.2000	7471B	PPB	21 Jun 2024 11:45:12	S	Std
2.5 - 1	SO.5	14930	-	2.5000	7471B	PPB	21 Jun 2024 11:47:29	S	Std
5.0 - 1	SS	30772	-	5.0000	7471B	PPB	21 Jun 2024 11:49:46	S	Std
7.5 - 1	SO.5	45941	-	7.5000	7471B	PPB	21 Jun 2024 11:52:05	S	Std
10.0 - 1	SO	61093	-	10.0000	7471B	PPB	21 Jun 2024 11:54:23	S	Std
ICV99 - 1	ICV99	22848	3.7467	-	7471B	PPB	21 Jun 2024 11:57:14	U	SMPL
ICB99 - 1	ICB99	-206	-0.0123	-	7471B	PPB	21 Jun 2024 11:59:31	U	SMPL
CCV71 - 1	CCV71	30526	4.9986	-	7471B	PPB	21 Jun 2024 12:01:49	U	SMPL
CCB71 - 1	CCB71	-237	-0.0173	-	7471B	PPB	21 Jun 2024 12:04:05	U	SMPL
CRA - 1	CRA	962	0.1782	-	7471B	PPB	21 Jun 2024 12:06:24	U	SMPL
HighStd - 1	HighStd	60657	9.9115	-	7471B	PPB	21 Jun 2024 12:08:42	U	SMPL
ChkStd - 1	ChkStd	42244	6.9092	-	7471B	PPB	21 Jun 2024 12:10:58	U	SMPL
PB161616BL - 1	PBS	-224	-0.0152	-	7471B	PPB	21 Jun 2024 12:13:22	U	SMPL
PB161616BS - 1	LLLLLLSS	20811	3.4146	-	7471B	PPB	21 Jun 2024 12:15:43	U	SMPL
P2930-01 - 1	WCS-TPE	3766	0.6354	-	7471B	PPB	21 Jun 2024 12:18:00	U	SMPL
P2940-01 - 1	SU-03-061824	931	0.1731	-	7471B	PPB	21 Jun 2024 12:20:19	U	SMPL
P2940-01DUP - 1	SU-03-061824DUP	1153	0.2093	-	7471B	PPB	21 Jun 2024 12:22:37	U	SMPL
P2940-01MS - 1	SU-03-061824MS	22873	3.7508	-	7471B	PPB	21 Jun 2024 12:24:54	U	SMPL
P2940-01MSD - 1	SU-03-061824MSD	21809	3.5773	-	7471B	PPB	21 Jun 2024 12:27:10	U	SMPL
CCV72 - 1	CCV72	29468	4.8261	-	7471B	PPB	21 Jun 2024 12:29:30	U	SMPL
CCB72 - 1	CCB72	-162	-0.0051	-	7471B	PPB	21 Jun 2024 12:31:48	U	SMPL
P2943-01 - 1	VNJ-212	3008	0.5118	-	7471B	PPB	21 Jun 2024 12:34:06	U	SMPL
P2943-03 - 1	72-11978	3746	0.6321	-	7471B	PPB	21 Jun 2024 12:36:22	U	SMPL
P2952-01 - 1	72-11944	5726	0.9549	-	7471B	PPB	21 Jun 2024 12:38:38	U	SMPL
P2953-01 - 1	SOM-24-00054	2489	0.4271	-	7471B	PPB	21 Jun 2024 12:40:56	U	SMPL
P2953-02 - 1	COMP-1	82	0.0347	-	7471B	PPB	21 Jun 2024 12:43:13	U	SMPL
P2953-03 - 1	COMP-2	49	0.0293	-	7471B	PPB	21 Jun 2024 12:45:31	U	SMPL
P2955-01 - 1	VK-1	6831	1.1351	-	7471B	PPB	21 Jun 2024 12:47:47	U	SMPL
P2955-05 - 1	TP-1	9169	1.5163	-	7471B	PPB	21 Jun 2024 12:50:04	U	SMPL
P2976-01 - 1	MH-K-COMP	456	0.0957	-	7471B	PPB	21 Jun 2024 12:52:23	U	SMPL
P2995-01 - 1	HR-02-06202024	2866	0.4886	-	7471B	PPB	21 Jun 2024 12:54:41	U	SMPL
CCV73 - 1	CCV73	29665	4.8582	-	7471B	PPB	21 Jun 2024 12:56:57	U	SMPL
CCB73 - 1	CCB73	-145	-0.0023	-	7471B	PPB	21 Jun 2024 12:59:13	U	SMPL
P2997-01 - 1	TR-06-062024	2812	0.4798	-	7471B	PPB	21 Jun 2024 13:01:32	U	SMPL
P2998-01 - 1	WCS-TPL	1006	0.1853	-	7471B	PPB	21 Jun 2024 13:03:48	U	SMPL
P2940-01LX5 - 1		67	0.0322	-	7471B	PPB	21 Jun 2024 13:06:05	U	SMPL
P2940-01A - 1		26764	4.3852	-	7471B	PPB	21 Jun 2024 13:08:21	U	SMPL
CCV74 - 1	CCV74	29739	4.8703	-	7471B	PPB	21 Jun 2024 13:10:37	U	SMPL
CCB74 - 1	CCB74	-134	-0.0005	-	7471B	PPB	21 Jun 2024 13:12:56	U	SMPL