

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB131350

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3085

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.7	06/21/2024	08:00
2	P2993-01		1	13	100.00	>212.0	>212.0	06/21/2024	08:30
3	P2993-01DUP		1	13	100.00	>212.0	>212.0	06/21/2024	09:00
4	P2993-02		1	11	100.00	>212.0	>212.0	06/21/2024	09:30
5	P2993-03		1	14	100.00	>212.0	>212.0	06/21/2024	10:00
6	P2993-04		1	12	100.00	>212.0	>212.0	06/21/2024	10:30
7	P2993-05		1	14	100.00	>212.0	>212.0	06/21/2024	11:00
8	P2996-04		1	12	100.00	>212.0	>212.0	06/21/2024	11:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

66131350

WorkList Name : FP-620

WorkList ID : 181254

Department : Wet-Chemistry

Date : 06-20-2024 15:38:48

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2993-01 D	619	Water	Flash Point	Cool 4 deg C	PSEG03	M21	06/20/2024	1010B
P2993-02 D	56A	Water	Flash Point	Cool 4 deg C	PSEG03	M21	06/20/2024	1010B
P2993-03 D	528A	Water	Flash Point	Cool 4 deg C	PSEG03	M21	06/20/2024	1010B
P2993-04 B	515G	Water	Flash Point	Cool 4 deg C	PSEG03	M21	06/20/2024	1010B
P2993-05 F	453	Water	Flash Point	Cool 4 deg C	PSEG03	M21	06/20/2024	1010B
P2996-04 B	COMP-33710-33711-34342	Water	Flash Point	Cool 4 deg C	PSEG03	M21	06/20/2024	1010B

Date/Time 06/21/2024 07:50
 Raw Sample Received by: RM CWC
 Raw Sample Relinquished by: P604

Date/Time 06/21/2024 12:10
 Raw Sample Received by: P604
 Raw Sample Relinquished by: RM CWC