

LB131361


SFAM01.1

INSTRUMENT ID : CY1

Linear

 μ Abs.:

53918



A= 0.0000e+000

B= 1.8828e-004 *5142*C= 2.8714e-002 *intercept*

Rho= 0.9994069

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>0/1)</i>
0.00	0.000	0.060	0.060	164	0.000	164					1
0.05	0.050					442	0				7
0.20	0.200	0.213	0.013	979	0.0 %	979					3
2.50	2.500	2.576	0.076	13530	0.0 %	13530					3
5.00	5.000	4.850	-0.150	25607	0.0 %	25607					2
7.50	7.500	7.320	-0.180	38727	0.0 %	38727					2
10.0	10.000	10.181	0.181	53918	0.0 %	53918					2

LB131361
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	164	-		0 SFAM01.1	PPB	6/21/2024 17:32	Std
	0.2 S01	979	-		0.2 SFAM01.1	PPB	6/21/2024 17:34	Std
	2.5 S02	13530	-		2.5 SFAM01.1	PPB	6/21/2024 17:37	Std
	5 S03	25607	-		5 SFAM01.1	PPB	6/21/2024 17:39	Std
	7.5 S04	38727	-		7.5 SFAM01.1	PPB	6/21/2024 17:41	Std
	10 S05	53918	-		10 SFAM01.1	PPB	6/21/2024 17:46	Std
ICV001	ICV001	19342	3.6705 -		SFAM01.1	PPB	6/21/2024 17:49	SMPL
ICB001	ICB001	102	0.0479 -		SFAM01.1	PPB	6/21/2024 17:51	SMPL
CCV075	CCV075	24927	4.7221 -		SFAM01.1	PPB	6/21/2024 17:54	SMPL
CCB075	CCB075	20	0.0325 -		SFAM01.1	PPB	6/21/2024 17:56	SMPL
PB161644BL	PBS644	33	0.0349 -		SFAM01.1	PPB	6/21/2024 17:58	SMPL
P2724-01	MA4D86	416313	HIGH -		SFAM01.1	PPB	6/21/2024 18:00	
P2724-02	MA4BP6	16962	3.2224 -		SFAM01.1	PPB	6/21/2024 18:02	SMPL
P2724-03	MA4BP6D	14951	2.8438 -		SFAM01.1	PPB	6/21/2024 18:05	SMPL
P2724-04	MA4BP6S	27371	5.1822 -		SFAM01.1	PPB	6/21/2024 18:07	SMPL
P2724-05	MA4BQ7	9436	1.8054 -		SFAM01.1	PPB	6/21/2024 18:09	SMPL
P2724-06	MA4CX9	1847	0.3765 -		SFAM01.1	PPB	6/21/2024 18:12	SMPL
P2724-07	MA4CY0	348	0.0942 -		SFAM01.1	PPB	6/21/2024 18:14	SMPL
P2724-08	MA4CY2	338	0.0924 -		SFAM01.1	PPB	6/21/2024 18:16	SMPL
P2724-09	MA4C86	6254	1.2062 -		SFAM01.1	PPB	6/21/2024 18:18	SMPL
P2724-10	MA4C87	6833	1.3153 -		SFAM01.1	PPB	6/21/2024 18:21	SMPL
P2724-11	MA4C88	1893	0.3851 -		SFAM01.1	PPB	6/21/2024 18:23	SMPL
P2724-12	MA4C90	4585	0.892 -		SFAM01.1	PPB	6/21/2024 18:25	SMPL
P2724-13	MA4CA1	1337	0.2805 -		SFAM01.1	PPB	6/21/2024 18:27	SMPL
P2724-14	MA4D56	930	0.2038 -		SFAM01.1	PPB	6/21/2024 18:30	SMPL
P2724-15	MA4D58	1121	0.2398 -		SFAM01.1	PPB	6/21/2024 18:32	SMPL
P2724-16	MA4D59	918	0.2016 -		SFAM01.1	PPB	6/21/2024 18:34	SMPL
P2724-17	MA4C93	8276	1.587 -		SFAM01.1	PPB	6/21/2024 18:36	SMPL
P2724-18	MA4C95	6525	1.2573 -		SFAM01.1	PPB	6/21/2024 18:39	SMPL
P2724-19	MA4C96	1586	0.3273 -		SFAM01.1	PPB	6/21/2024 18:41	SMPL
CCV076	CCV076	23483	4.4502 -		SFAM01.1	PPB	6/21/2024 18:43	SMPL
CCB076	CCB076	-116	0.0069 -		SFAM01.1	PPB	6/21/2024 18:46	SMPL
P2724-20	MA4C97	1285	0.2707 -		SFAM01.1	PPB	6/21/2024 18:48	SMPL
P2724-21	MA4D46	6346	1.2236 -		SFAM01.1	PPB	6/21/2024 18:50	SMPL
P2724-22	MA4D47	1810	0.3695 -		SFAM01.1	PPB	6/21/2024 18:53	SMPL
P2724-23	MA4D48	1500	0.3111 -		SFAM01.1	PPB	6/21/2024 18:55	SMPL
PB161645BL	PBS645	-8	0.0272 -		SFAM01.1	PPB	6/21/2024 18:57	SMPL
P2755-01	MA4B87	10037	1.9185 -		SFAM01.1	PPB	6/21/2024 18:59	SMPL
P2755-02	MA4B87D	9960	1.904 -		SFAM01.1	PPB	6/21/2024 19:02	SMPL
P2755-03	MA4B87S	25052	4.7456 -		SFAM01.1	PPB	6/21/2024 19:04	SMPL
P2755-04	MA4B88	7535	1.4474 -		SFAM01.1	PPB	6/21/2024 19:06	SMPL
P2755-05	MA4B89	430	0.1097 -		SFAM01.1	PPB	6/21/2024 19:08	SMPL
P2755-06	MA4B90	8423	1.6146 -		SFAM01.1	PPB	6/21/2024 19:11	SMPL
P2755-07	MA4B91	4852	0.9423 -		SFAM01.1	PPB	6/21/2024 19:13	SMPL
P2755-08	MA4BP5	4137	0.8076 -		SFAM01.1	PPB	6/21/2024 19:15	SMPL
P2755-09	MA4BP9	11496	2.1932 -		SFAM01.1	PPB	6/21/2024 19:18	SMPL

LB131361

INSTRUMENT ID : CV1

P2755-10	MA4BM5	12055	2.2985 -	SFAM01.1	PPB	6/21/2024 19:20 SMPL
P2755-11	MA4BM6	10851	2.0718 -	SFAM01.1	PPB	6/21/2024 19:22 SMPL
P2755-12	MA4BP0	9332	1.7858 -	SFAM01.1	PPB	6/21/2024 19:24 SMPL
P2755-13	MA4BP2	14585	2.7748 -	SFAM01.1	PPB	6/21/2024 19:27 SMPL
P2755-14	MA4BP3	14543	2.7669 -	SFAM01.1	PPB	6/21/2024 19:29 SMPL
P2755-15	MA4BP4	11635	2.2194 -	SFAM01.1	PPB	6/21/2024 19:31 SMPL
CCV077	CCV077	24788	4.6959 -	SFAM01.1	PPB	6/21/2024 19:34 SMPL
CCB077	CCB077	-183	-0.0057 -	SFAM01.1	PPB	6/21/2024 19:36 SMPL
P2755-16	MA4BP7	1348	0.2825 -	SFAM01.1	PPB	6/21/2024 19:38 SMPL
P2755-17	MA4BP8	10188	1.947 -	SFAM01.1	PPB	6/21/2024 19:40 SMPL
P2755-18	MA4D87	268384	50.5613 -	SFAM01.1	PPB	6/21/2024 19:43 SMPL
P2755-19	MA4B92	9753	1.8651 -	SFAM01.1	PPB	6/21/2024 19:45 SMPL
P2755-20	MA4BQ0	17254	3.2774 -	SFAM01.1	PPB	6/21/2024 19:48 SMPL
P2755-21	MA4BQ5	7366	1.4156 -	SFAM01.1	PPB	6/21/2024 19:50 SMPL
P2755-22	MA4BQ1	4411	0.8592 -	SFAM01.1	PPB	6/21/2024 19:52 SMPL
P2755-23	MA4BQ4	13608	2.5909 -	SFAM01.1	PPB	6/21/2024 19:55 SMPL
CCV078	CCV078	25778	4.8823 -	SFAM01.1	PPB	6/21/2024 20:14 SMPL
CCB078	CCB078	-155	-0.0005 -	SFAM01.1	PPB	6/21/2024 20:17 SMPL
P2724-01DLX10	MA4D86	70021	13.2126 -	SFAM01.1	PPB	6/21/2024 20:19 SMPL
P2755-18DLX10	MA4D87	29841	5.6473 -	SFAM01.1	PPB	6/21/2024 20:33 SMPL
P2724-01DL2X25	MA4D86	26523	5.0226 -	SFAM01.1	PPB	6/21/2024 20:40 SMPL
CCV079	CCV079	23864	4.5219 -	SFAM01.1	PPB	6/21/2024 20:47 SMPL
CCB079	CCB079	77	0.0432 -	SFAM01.1	PPB	6/21/2024 20:50 SMPL