

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB131382

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3085

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.4	06/24/2024	11:00
2	P2994-03		1	4	7.00	44.6	44.6	06/24/2024	11:30
3	P2994-03DUP		1	4	7.00	44.6	44.6	06/24/2024	12:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

66131382

WorkList Name : FP-S6-20

WorkList ID : 181256

Department : Wet-Chemistry

Date : 06-20-2024 16:16:22

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2994-03	32518	Solid	Flash Point	Cool 4 deg C	PSEG03	M21	06/20/2024	1010B

Date/Time 06/24/2024 10:45
Raw Sample Received by: RM CWC
Raw Sample Relinquished by: JDC

Date/Time 06/24/2024 12:40
Raw Sample Received by: RM CWC
Raw Sample Relinquished by: RM CWC