

LB131383


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

47341



A= 0.0000e+000

B= 2.0983e-004

C= -1.4895e-002

Rho= 0.9997436

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.005	0.005	96	0.000	96					
0.05	0.050					442	0				
0.20	0.200	0.234	0.034	1185	0.0 %	1185					
2.50	2.500	2.443	-0.057	11712	0.0 %	11712					
5.00	5.000	4.936	-0.064	23594	0.0 %	23594					
7.50	7.500	7.664	0.164	36593	0.0 %	36593					
10.0	10.000	9.919	-0.081	47341	0.0 %	47341					

LB131383

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	96	-		0 SFAM01.1	PPB	6/24/2024 13:01	Std
	0.2 S01	1185	-		0.2 SFAM01.1	PPB	6/24/2024 13:04	Std
	2.5 S02	11712	-		2.5 SFAM01.1	PPB	6/24/2024 13:06	Std
	5 S03	23594	-		5 SFAM01.1	PPB	6/24/2024 13:08	Std
	7.5 S04	36593	-		7.5 SFAM01.1	PPB	6/24/2024 13:10	Std
	10 S05	47341	-		10 SFAM01.1	PPB	6/24/2024 13:15	Std
ICV002	ICV002	19157	4.0049 -		SFAM01.1	PPB	6/24/2024 13:18	SMPL
ICB002	ICB002	76	0.0011 -		SFAM01.1	PPB	6/24/2024 13:23	SMPL
CCV079	CCV079	24143	5.0511 -		SFAM01.1	PPB	6/24/2024 13:25	SMPL
CCB079	CCB079	90	0.004 -		SFAM01.1	PPB	6/24/2024 13:30	SMPL
PB161655BL	PBW655	11	-0.0126 -		SFAM01.1	PPB	6/24/2024 13:32	SMPL
P2409-01	MDL-WATER	367	0.0621 -		SFAM01.1	PPB	6/24/2024 13:37	SMPL
P2781-20	MA4D82	11	-0.0126 -		SFAM01.1	PPB	6/24/2024 13:39	SMPL
P3018-01	E08C0	74	0.0006 -		SFAM01.1	PPB	6/24/2024 13:42	SMPL
P3018-02	E08C0D	45	-0.0055 -		SFAM01.1	PPB	6/24/2024 13:44	SMPL
P3018-03	E08C0S	4512	0.9319 -		SFAM01.1	PPB	6/24/2024 13:46	SMPL
P3019-01	E08C1	-9	-0.0168 -		SFAM01.1	PPB	6/24/2024 13:49	SMPL
P3019-02	E08C1D	51	-0.0042 -		SFAM01.1	PPB	6/24/2024 13:51	SMPL
P3019-03	E08C1S	4986	1.0313 -		SFAM01.1	PPB	6/24/2024 13:53	SMPL
P3038-01	E08C4	1075	0.2107 -		SFAM01.1	PPB	6/24/2024 13:55	SMPL
P3038-02	E08C4D	1060	0.2075 -		SFAM01.1	PPB	6/24/2024 13:58	SMPL
P3038-03	E08C4S	5914	1.2261 -		SFAM01.1	PPB	6/24/2024 14:00	SMPL
P3039-01	E08C5	1004	0.1958 -		SFAM01.1	PPB	6/24/2024 14:02	SMPL
P3039-02	E08C5D	933	0.1809 -		SFAM01.1	PPB	6/24/2024 14:07	SMPL
P3039-03	E08C5S	6701	1.3912 -		SFAM01.1	PPB	6/24/2024 14:09	SMPL
CCV080	CCV080	23446	4.9049 -		SFAM01.1	PPB	6/24/2024 14:11	SMPL
CCB080	CCB080	65	-0.0013 -		SFAM01.1	PPB	6/24/2024 14:16	SMPL