

LB131395

7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

39279

A= 0.0000e+000

B= 2.5201e-004

C= -1.5103e-002

Rho= 0.9996584

Accept=Accepted

Max. Conc: 10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/00
0.0	0.000	-0.018	-0.018	-12	0.000	-12	0				
0.2	0.200	0.200	0.000	855	0.0 %	855					
2.5	2.500	2.399	-0.101	9578	0.0 %	9578					
5.0	5.000	5.162	0.162	20543	0.0 %	20543					
7.5	7.500	7.574	0.074	30113	0.0 %	30113					
10.0	10.000	9.884	-0.116	39279	0.0 %	39279					

0/00
1
-4
3
1

LB131395 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 25 Jun 2024 07:59:46

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-12	-	0.0000	7470A	PPB	25 Jun 2024 08:09:19	S	Std
0.2 - 1	50.2	855	-	0.2000	7470A	PPB	25 Jun 2024 08:11:36	S	Std
2.5 - 1	52.5	9578	-	2.5000	7470A	PPB	25 Jun 2024 08:13:53	S	Std
5.0 - 1	55	20543	-	5.0000	7470A	PPB	25 Jun 2024 08:16:10	S	Std
7.5 - 1	57.5	30113	-	7.5000	7470A	PPB	25 Jun 2024 08:18:27	S	Std
10.0 - 1	50	39279	-	10.0000	7470A	PPB	25 Jun 2024 08:23:21	S	Std
ICV03 - 1	ICV03	15942	4.0024	-	7470A	PPB	25 Jun 2024 08:26:08	U	SMPL
ICB03 - 1	ICB03	-49	-0.0275	-	7470A	PPB	25 Jun 2024 08:31:27	U	SMPL
CCV81 - 1	CCV81	20510	5.1536	-	7470A	PPB	25 Jun 2024 08:33:42	U	SMPL
CCB81 - 1	CCB81	-148	-0.0524	-	7470A	PPB	25 Jun 2024 08:35:57	U	SMPL
CRA - 1	CRA	889	0.2089	-	7470A	PPB	25 Jun 2024 08:38:13	U	SMPL
HighStd - 1	HighStd	39924	10.0461	-	7470A	PPB	25 Jun 2024 08:40:29	U	SMPL
ChkStd - 1	ChkStd	28080	7.0613	-	7470A	PPB	25 Jun 2024 08:42:46	U	SMPL
PB161661BL - 1	PBW	-98	-0.0398	-	7470A	PPB	25 Jun 2024 08:47:40	U	SMPL
PB161661BS - 1	LCSW	14336	3.5977	-	7470A	PPB	25 Jun 2024 08:49:55	U	SMPL
P2998-04 - 1	WCS-TPL	-249	-0.0779	-	7470A	PPB	25 Jun 2024 08:52:11	U	SMPL
P3008-03 - 1	WC-WS-6-21-24	29	-0.0078	-	7470A	PPB	25 Jun 2024 08:54:29	U	SMPL
P3009-03 - 1	VNJ-204	-63	-0.0310	-	7470A	PPB	25 Jun 2024 08:56:46	U	SMPL
P3021-03 - 1	ARS20-008	115	0.0139	-	7470A	PPB	25 Jun 2024 08:59:03	U	SMPL
P3021-06 - 1	ARS20-0018	167	0.0270	-	7470A	PPB	25 Jun 2024 09:01:19	U	SMPL
CCV82 - 1	CCV82	20159	5.0651	-	7470A	PPB	25 Jun 2024 09:03:34	U	SMPL
CCB82 - 1	CCB82	-229	-0.0728	-	7470A	PPB	25 Jun 2024 09:05:50	U	SMPL
P3021-09 - 1	ARS20-006	-70	-0.0327	-	7470A	PPB	25 Jun 2024 09:08:07	U	SMPL
P3021-12 - 1	ARS20-0015	-52	-0.0282	-	7470A	PPB	25 Jun 2024 09:10:23	U	SMPL
P3021-12DUP - 1	ARS20-0015DUP	-50	-0.0277	-	7470A	PPB	25 Jun 2024 09:12:38	U	SMPL
P3021-12MS - 1	ARS20-0015MS	13415	3.3656	-	7470A	PPB	25 Jun 2024 09:14:54	U	SMPL
P3021-12MSD - 1	ARS20-0015MSD	15623	3.9220	-	7470A	PPB	25 Jun 2024 09:17:10	U	SMPL
PB161643TB - 1	PB161643TB	226	0.0419	-	7470A	PPB	25 Jun 2024 09:19:27	U	SMPL
P3021-12LX5 - 1		7	-0.0133	-	7470A	PPB	25 Jun 2024 09:21:46	U	SMPL
P3021-12A - 1		14935	3.7487	-	7470A	PPB	25 Jun 2024 09:24:03	U	SMPL
CCV83 - 1	CCV83	20132	5.0583	-	7470A	PPB	25 Jun 2024 09:26:20	U	SMPL
CCB83 - 1	CCB83	-165	-0.0567	-	7470A	PPB	25 Jun 2024 09:28:37	U	SMPL