

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB131409

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	06/26/2024	08:00
2	CAL2	1	Water	NA	NA	20.3	6.98	06/26/2024	08:01
3	CAL3	1	Water	NA	NA	20.2	10.02	06/26/2024	08:05
4	ICV	1	Water	NA	NA	20.4	7.00	06/26/2024	08:10
5	CCV1	1	Water	NA	NA	20.3	2.01	06/26/2024	08:15
6	P3051-01	1	Water	NA	NA	23.2	5.52	06/26/2024	08:25
7	P3069-04	1	Water	NA	NA	23.7	7.24	06/26/2024	08:30
8	P3069-04DUP	1	Water	NA	NA	23.8	7.25	06/26/2024	08:31
9	CCV2	1	Water	NA	NA	20.4	12.02	06/26/2024	08:35

WORKLIST(Hardcopy Internal Chain)

VB 131409

WorkList Name : PH P3069 WorkList ID : 181398 Department : Wet-Chemistry Date : 06-26-2024 07:25:19

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3051-01	TRE-1299	Water	pH	Cool 4 deg C	PSEG03	M21	06/24/2024	9040C
P3069-04	A1285	Water	pH	Cool 4 deg C	PSEG03	M21	06/25/2024	9040C

Date/Time 06/26/24 07:30

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 06/26/24

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

10:00

[Signature]