


LB131416

7471B

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0)
0.0	0.000	-0.010	-0.010	53	0.000	53					-
0.2	0.200	0.214	0.014	1213	0.0 %	1213					7
2.5	2.500	2.364	-0.136	12361	0.0 %	12361					5
5.0	5.000	5.074	0.074	26417	0.0 %	26417					1
7.5	7.500	7.827	0.327	40693	0.0 %	40693					4
10.0	10.000	9.732	-0.268	50573	0.0 %	50573					-3

LB131416 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 26 Jun 2024 11:43:50

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	53	-	0.0000	7471B	PPB	26 Jun 2024 11:50:34	S	Std
0.2 - 1	50-2	1213	-	0.2000	7471B	PPB	26 Jun 2024 11:52:51	S	Std
2.5 - 1	50.5	12361	-	2.5000	7471B	PPB	26 Jun 2024 11:55:08	S	Std
5.0 - 1	55	26417	-	5.0000	7471B	PPB	26 Jun 2024 11:57:26	S	Std
7.5 - 1	57.5	40693	-	7.5000	7471B	PPB	26 Jun 2024 12:02:25	S	Std
10.0 - 1	510	50573	-	10.0000	7471B	PPB	26 Jun 2024 12:10:01	S	Std
ICV06 - 1	ICV06	22408	4.3008	-	7471B	PPB	26 Jun 2024 12:12:57	U	SMPL
ICB06 - 1	ICB06	-302	-0.0781	-	7471B	PPB	26 Jun 2024 12:15:13	U	SMPL
CCV92 - 1	CCV92	26491	5.0881	-	7471B	PPB	26 Jun 2024 12:17:33	U	SMPL
CCB92 - 1	CCB92	32	-0.0137	-	7471B	PPB	26 Jun 2024 12:27:37	U	SMPL
CRA - 1	CRA	1212	0.2138	-	7471B	PPB	26 Jun 2024 12:29:51	U	SMPL
HighStd - 1	HighStd	50398	9.6979	-	7471B	PPB	26 Jun 2024 12:32:07	U	SMPL
ChkStd - 1	ChkStd	36929	7.1008	-	7471B	PPB	26 Jun 2024 12:34:22	U	SMPL
PB161742BL - 1	PBS	-380	-0.0932	-	7471B	PPB	26 Jun 2024 12:36:44	U	SMPL
PB161742BS - 1	LCSS	19443	3.7291	-	7471B	PPB	26 Jun 2024 12:39:06	U	SMPL
P2656-11 - 1	SB-21(0-2)	622769	HIGH	-	7471B	PPB	26 Jun 2024 12:41:21	X	
P2656-12 - 1	SB-22(0-2)	1274233	HIGH	-	7471B	PPB	26 Jun 2024 12:43:12	X	
P2656-13 - 1	SB-23(0-2)	675941	HIGH	-	7471B	PPB	26 Jun 2024 12:45:35	X	
P2656-14 - 1	SB-24(0-3)	139625	26.9028	-	7471B	PPB	26 Jun 2024 12:48:31	U	SMPL
P2656-15 - 1	SB-25(0-2)	1351279	HIGH	-	7471B	PPB	26 Jun 2024 12:51:12	X	
CCV93 - 1	CCV93	25233	4.8456	-	7471B	PPB	26 Jun 2024 12:53:21	U	SMPL
CCB93 - 1	CCB93	-158	-0.0504	-	7471B	PPB	26 Jun 2024 12:59:51	U	SMPL
P2656-16 - 1	SB-26(0-2)	797300	HIGH	-	7471B	PPB	26 Jun 2024 13:02:06	X	
P3045-01 - 1	6AB	30735	5.9065	-	7471B	PPB	26 Jun 2024 13:03:53	U	SMPL
P3045-01DUP - 1	6ABDUP	30453	5.8521	-	7471B	PPB	26 Jun 2024 13:06:54	U	SMPL
P3045-01MS - 1	6ABMS	49076	9.4430	-	7471B	PPB	26 Jun 2024 13:09:16	U	SMPL
P3045-01MSD - 1	6ABMSD	49023	9.4328	-	7471B	PPB	26 Jun 2024 13:11:37	U	SMPL
P3045-03 - 1	7AB	4303	0.8098	-	7471B	PPB	26 Jun 2024 13:13:58	U	SMPL
P3045-05 - 1	8AB	18005	3.4519	-	7471B	PPB	26 Jun 2024 13:16:17	U	SMPL
P3045-07 - 1	9AB	3632	0.6804	-	7471B	PPB	26 Jun 2024 13:18:37	U	SMPL
P3045-09 - 1	10AB	3195	0.5962	-	7471B	PPB	26 Jun 2024 13:20:56	U	SMPL
P3047-01 - 1	EO-01-062424	18538	3.5546	-	7471B	PPB	26 Jun 2024 13:23:13	U	SMPL
CCV94 - 1	CCV94	26397	5.0700	-	7471B	PPB	26 Jun 2024 13:25:29	U	SMPL
CCB94 - 1	CCB94	-157	-0.0502	-	7471B	PPB	26 Jun 2024 13:27:48	U	SMPL
P3049-01 - 1	SHOP-RAG-DRUM	1619	0.2923	-	7471B	PPB	26 Jun 2024 13:30:07	U	SMPL
P3053-01 - 1	1A	3085	0.5750	-	7471B	PPB	26 Jun 2024 13:32:22	U	SMPL
P3045-01LX5 - 1		6700	1.2720	-	7471B	PPB	26 Jun 2024 13:34:38	U	SMPL
P2656-11DLX10 - 1		95821	18.4564	-	7471B	PPB	26 Jun 2024 13:39:13	U	SMPL
P2656-12DLX10 - 1		368287	70.9937	-	7471B	PPB	26 Jun 2024 13:41:34	U	SMPL
P2656-13DLX10 - 1		98983	19.0661	-	7471B	PPB	26 Jun 2024 13:43:57	U	SMPL
P2656-14DLX10 - 1		34699	6.6708	-	7471B	PPB	26 Jun 2024 13:46:34	U	SMPL
P2656-15DLX10 - 1		465454	HIGH	-	7471B	PPB	26 Jun 2024 13:49:03	X	
P2656-16DLX10 - 1		153048	29.4910	-	7471B	PPB	26 Jun 2024 13:50:55	U	SMPL
CCV95 - 1	CCV95	26620	5.1130	-	7471B	PPB	26 Jun 2024 13:53:28	U	SMPL
CCB95 - 1	CCB95	-309	-0.0795	-	7471B	PPB	26 Jun 2024 13:55:59	U	SMPL
P3045-01A - 1		47080	9.0581	-	7471B	PPB	26 Jun 2024 13:58:19	U	SMPL
P2656-11DL2X50 - 1		18955	3.6350	-	7471B	PPB	26 Jun 2024 14:03:47	U	SMPL
P2656-1DL2X100 - 1 P2656-12DL2X100		35827	6.8883	-	7471B	PPB	26 Jun 2024 14:06:02	U	SMPL
P2656-13DL2X50 - 1		23497	4.5108	-	7471B	PPB	26 Jun 2024 14:08:20	U	SMPL
P2656-15DL2X500 - 1		16968	3.2519	-	7471B	PPB	26 Jun 2024 14:19:23	U	SMPL
P2656-16DL2X50 - 1		36020	6.9255	-	7471B	PPB	26 Jun 2024 14:21:39	U	SMPL
CCV97 - 1	CCV97	26142	5.0208	-	7471B	PPB	26 Jun 2024 14:23:58	U	SMPL
CCB97 - 1	CCB97	-231	-0.0644	-	7471B	PPB	26 Jun 2024 14:26:18	U	SMPL

MB

08/06/2024