

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 6/27/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 17:00
In Date: 06/26/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:00
Out Date: 06/27/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB131417

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
P2857-04	ETGI-355	12	1.00	1.00	2.00	2.00	100.0	concrete sample, 100% solids
P3064-01	HD-02-062524	1	1.17	8.38	9.55	8.98	93.2	
P3064-02	HD-02-062524-E2	2	1.16	8.46	9.62	8.55	87.4	
P3076-01	7-N-1	3	1.17	8.60	9.77	8.98	90.8	
P3076-04	7-W-1	4	1.16	8.72	9.88	9.05	90.5	
P3076-07	7-E-1	5	1.16	8.82	9.98	9.15	90.6	
P3076-10	7-S-1	6	1.17	8.81	9.98	9.12	90.2	
P3076-13	7-B-1	7	1.17	8.51	9.68	8.79	89.5	
P3077-01	1-N-1	8	1.17	8.57	9.74	8.47	85.2	
P3077-04	1-W-1	9	1.16	8.68	9.84	8.98	90.1	
P3077-07	1-S-1	10	1.17	8.47	9.64	8.7	88.9	
P3077-10	1-B-1	11	1.17	8.34	9.51	8.71	90.4	
P3086-01	PCB052924RB01	13	1.00	1.00	2.00	2.00	100.0	CAULK
P3086-02	PCB052924RB02	14	1.00	1.00	2.00	2.00	100.0	CAULK

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$