

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB131420

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3085

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.7	06/26/2024	10:10
2	P3051-01		1	14	100.00	>212.0	>212.0	06/26/2024	10:40
3	P3051-01DUP		1	14	100.00	>212.0	>212.0	06/26/2024	11:10
4	P3069-02		1	12	100.00	>212.0	>212.0	06/26/2024	11:40
5	P3069-03		1	12	100.00	>212.0	>212.0	06/26/2024	12:10
6	P3069-04		1	13	100.00	>212.0	>212.0	06/26/2024	13:10
7	P3072-01		1	14	100.00	>212.0	>212.0	06/26/2024	13:40
8	P3072-02		1	12	100.00	>212.0	>212.0	06/26/2024	14:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp-6-26

WorkList ID : 181404

Department : Wet-Chemistry

Date : 06-26-2024 08:00:21

66131420

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3051-01	TRE-1299	Water	Flash Point	Cool 4 deg C	PSEG03	M21	06/24/2024	1010B
P3069-02	5302	Water	Flash Point	Cool 4 deg C	PSEG03	M21	06/25/2024	1010B
P3069-03	611	Water	Flash Point	Cool 4 deg C	PSEG03	M21	06/25/2024	1010B
P3069-04	A1285	Water	Flash Point	Cool 4 deg C	PSEG03	M21	06/25/2024	1010B
P3072-01	44-W	Water	Flash Point	Cool 4 deg C	PSEG03	M21	06/25/2024	1010B
P3072-02	1287	Water	Flash Point	Cool 4 deg C	PSEG03	M21	06/25/2024	1010B

Date/Time 06/26/2024 10:00
 Raw Sample Received by: RM CWG
 Raw Sample Relinquished by: PCG

Date/Time 06/26/2024 14:58
 Raw Sample Received by: ARCG
 Raw Sample Relinquished by: RM CWG