

LB131427


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

35162



A= 0.0000e+000

B= 2.7334e-004

C= 4.5292e-002

Rho= 0.9974989

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	0.039	0.039	-23	0.000	-23	0				0/01
0.2	0.200	0.252	0.052	756	0.0 %	756					26
2.5	2.500	2.178	-0.322	7803	0.0 %	7803					-13
5.0	5.000	5.195	0.195	18839	0.0 %	18839					4
7.5	7.500	7.880	0.380	28663	0.0 %	28663					5
10.0	10.000	9.656	-0.344	35162	0.0 %	35162					-3

LB131427 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 27 Jun 2024 12:46:46

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-23	-	0.0000	7470A	PPB	27 Jun 2024 13:04:24	S	Std
0.2 - 1	502	756	-	0.2000	7470A	PPB	27 Jun 2024 13:06:42	S	Std
2.5 - 1	525	7803	-	2.5000	7470A	PPB	27 Jun 2024 13:08:59	S	Std
5.0 - 1	55	18839	-	5.0000	7470A	PPB	27 Jun 2024 13:13:59	S	Std
7.5 - 1	575	28663	-	7.5000	7470A	PPB	27 Jun 2024 13:16:15	S	Std
10.0 - 1	510	35162	-	10.0000	7470A	PPB	27 Jun 2024 13:23:46	S	Std
ICV08 - 1	ICV08	13894	3.8430	-	7470A	PPB	27 Jun 2024 13:28:31	U	SMPL
ICB08 - 1	ICB08	-228	-0.0170	-	7470A	PPB	27 Jun 2024 13:30:45	U	SMPL
CCV05 - 1	CCV05	19192	5.2912	-	7470A	PPB	27 Jun 2024 13:33:02	U	SMPL
CCB05 - 1	CCB05	-267	-0.0277	-	7470A	PPB	27 Jun 2024 13:35:17	U	SMPL
CRA - 1	CRA	546	0.1945	-	7470A	PPB	27 Jun 2024 13:37:34	U	SMPL
HighStd - 1	HighStd	34640	9.5137	-	7470A	PPB	27 Jun 2024 13:39:50	U	SMPL
ChkStd - 1	ChkStd	27869	7.6629	-	7470A	PPB	27 Jun 2024 13:44:45	U	SMPL
PB161761BL - 1	PBW	-286	-0.0329	-	7470A	PPB	27 Jun 2024 13:47:00	U	SMPL
PB161761BS - 1	LCSW	16256	4.4886	-	7470A	PPB	27 Jun 2024 13:49:19	U	SMPL
P3045-02 - 1	6AB	-278	-0.0307	-	7470A	PPB	27 Jun 2024 13:51:36	U	SMPL
P3045-02DUP - 1	6ABDUP	-8	0.0431	-	7470A	PPB	27 Jun 2024 13:53:54	U	SMPL
P3045-02MS - 1	6ABMS	14955	4.1330	-	7470A	PPB	27 Jun 2024 13:56:12	U	SMPL
P3045-02MSD - 1	6ABMSD	15005	4.1467	-	7470A	PPB	27 Jun 2024 13:58:29	U	SMPL
P3045-04 - 1	7AB	-191	-0.0069	-	7470A	PPB	27 Jun 2024 14:00:46	U	SMPL
CCV06 - 1	CCV06	18046	4.9779	-	7470A	PPB	27 Jun 2024 14:03:04	U	SMPL
CCB06 - 1	CCB06	-269	-0.0282	-	7470A	PPB	27 Jun 2024 14:05:19	U	SMPL
P3045-06 - 1	8AB	114	0.0765	-	7470A	PPB	27 Jun 2024 14:07:37	U	SMPL
P3045-08 - 1	9AB	-70	0.0262	-	7470A	PPB	27 Jun 2024 14:09:53	U	SMPL
P3045-10 - 1	10AB	20	0.0508	-	7470A	PPB	27 Jun 2024 14:12:09	U	SMPL
P3049-02 - 1	SHOR-RAG-DRUM	-57	0.0297	-	7470A	PPB	27 Jun 2024 14:14:25	U	SMPL
PB161804BL - 1	PBW	170	0.0918	-	7470A	PPB	27 Jun 2024 14:19:21	U	SMPL
PB161804BS - 1	LCSW	16232	4.4821	-	7470A	PPB	27 Jun 2024 14:21:37	U	SMPL
P2993-04 - 1	515G	1260	0.3897	-	7470A	PPB	27 Jun 2024 14:23:54	U	SMPL
P2993-05 - 1	453	2375	0.6945	-	7470A	PPB	27 Jun 2024 14:26:13	U	SMPL
P3051-01 - 1	TRE-1299	3382	0.9697	-	7470A	PPB	27 Jun 2024 14:28:31	U	SMPL
P3069-02 - 1	5302	747	0.2495	-	7470A	PPB	27 Jun 2024 14:30:49	U	SMPL
CCV07 - 1	CCV07	17802	4.9112	-	7470A	PPB	27 Jun 2024 14:33:07	U	SMPL
CCB07 - 1	CCB07	-166	-0.0001	-	7470A	PPB	27 Jun 2024 14:35:25	U	SMPL
P3069-03 - 1	611	329	0.1352	-	7470A	PPB	27 Jun 2024 14:37:42	U	SMPL
P3069-04 - 1	A1285	426	0.1617	-	7470A	PPB	27 Jun 2024 14:39:58	U	SMPL
P3072-02 - 1	1287	2496	0.7275	-	7470A	PPB	27 Jun 2024 14:42:13	U	SMPL
P3072-02DUP - 1	1287DUP	2033	0.6010	-	7470A	PPB	27 Jun 2024 14:44:29	U	SMPL
P3072-02MS - 1	1287MS	12018	3.3302	-	7470A	PPB	27 Jun 2024 14:46:46	U	SMPL
P3072-02MSD - 1	1287MSD	12229	3.3879	-	7470A	PPB	27 Jun 2024 14:49:03	U	SMPL
CCV08 - 1	CCV08	18080	4.9872	-	7470A	PPB	27 Jun 2024 14:51:22	U	SMPL
CCB08 - 1	CCB08	-190	-0.0066	-	7470A	PPB	27 Jun 2024 14:53:40	U	SMPL
PB161688TB - 1	PB161688TB	-13	0.0417	-	7470A	PPB	27 Jun 2024 14:56:00	U	SMPL
P3045-02LX5 - 1		0	0.0453	-	7470A	PPB	27 Jun 2024 14:58:16	U	SMPL
P3045-02A - 1		14645	4.0483	-	7470A	PPB	27 Jun 2024 15:00:32	U	SMPL
P3072-02LX6 - 1		244	0.1120	-	7470A	PPB	27 Jun 2024 15:02:47	U	SMPL
P3072-02A - 1		16526	4.5624	-	7470A	PPB	27 Jun 2024 15:05:05	U	SMPL
CCV0.9 - 1		17718	4.8883	-	7470A	PPB	27 Jun 2024 15:07:20	U	SMPL
CCB09 - 1	CCB09	-210	-0.0121	-	7470A	PPB	27 Jun 2024 15:09:38	U	SMPL