

LB13143 G


SFAM01.1

INSTRUMENT ID: CV2

Linear

 μ Abs.:

44553



A= 0.0000e+000

B= 2.2316e-004 slope

C= -4.5323e-002 y-intercept

Rho= 0.9995179

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	σ/σ)
0.00	0.000	-0.013	-0.013	144	0.000	144					=
0.05	0.050					442	0				12
0.20	0.200	0.224	0.024	1205	0.0 %	1205					-6
2.50	2.500	2.347	-0.153	10719	0.0 %	10719					4
5.00	5.000	5.206	0.206	23531	0.0 %	23531					-1
7.50	7.500	7.540	0.040	33989	0.0 %	33989					-1
10.0	10.000	9.897	-0.103	44553	0.0 %	44553					-1

LB131436

INSTRUMENT ID : CV1 -

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	144	-		0 SFAM01.1	PPB	6/27/2024 16:08	Std
	0.2 S01	1205	-		0.2 SFAM01.1	PPB	6/27/2024 16:10	Std
	2.5 S02	10719	-		2.5 SFAM01.1	PPB	6/27/2024 16:13	Std
	5 S03	23531	-		5 SFAM01.1	PPB	6/27/2024 16:17	Std
	7.5 S04	33989	-		7.5 SFAM01.1	PPB	6/27/2024 16:22	Std
	10 S05	44553	-		10 SFAM01.1	PPB	6/27/2024 16:27	Std
ICV009	ICV009	19419	4.2883 -		SFAM01.1	PPB	6/27/2024 16:31	SMPL
ICB009	ICB009	-24	-0.0507 -		SFAM01.1	PPB	6/27/2024 16:33	SMPL
CCV010	CCV010	23879	5.2836 -		SFAM01.1	PPB	6/27/2024 16:36	SMPL
CCB010	CCB010	-34	-0.0529 -		SFAM01.1	PPB	6/27/2024 16:38	SMPL
PB161808BL	PBS808	279	0.0169 -		SFAM01.1	PPB	6/27/2024 16:40	SMPL
P2845-01	MA4D90	393167	HIGH -		SFAM01.1	PPB	6/27/2024 16:42	
P2845-02	MA4C64	2054	0.4131 -		SFAM01.1	PPB	6/27/2024 16:44	SMPL
P2845-04	MA4C64S	14425	3.1738 -		SFAM01.1	PPB	6/27/2024 16:52	SMPL
P2845-05	MA4C46	18726	4.1336 -		SFAM01.1	PPB	6/27/2024 16:54	SMPL
P2845-06	MA4C47	7927	1.7237 -		SFAM01.1	PPB	6/27/2024 16:56	SMPL
P2845-07	MA4C48	4080	0.8652 -		SFAM01.1	PPB	6/27/2024 16:59	SMPL
P2845-08	MA4C53	1796	0.3555 -		SFAM01.1	PPB	6/27/2024 17:01	SMPL
P2845-09	MA4C57	21913	4.8448 -		SFAM01.1	PPB	6/27/2024 17:03	SMPL
P2845-10	MA4CN4	10674	2.3367 -		SFAM01.1	PPB	6/27/2024 17:06	SMPL
P2845-11	MA4CN5	1887	0.3758 -		SFAM01.1	PPB	6/27/2024 17:08	SMPL
P2845-12	MA4CN7	4668	0.9964 -		SFAM01.1	PPB	6/27/2024 17:10	SMPL
P2845-13	MA4C52	1908	0.3805 -		SFAM01.1	PPB	6/27/2024 17:12	SMPL
P2845-14	MA4C54	6331	1.3675 -		SFAM01.1	PPB	6/27/2024 17:15	SMPL
P2845-15	MA4C55	1227	0.2285 -		SFAM01.1	PPB	6/27/2024 17:17	SMPL
P2845-16	MA4C58	2118	0.4273 -		SFAM01.1	PPB	6/27/2024 17:19	SMPL
P2845-17	MA4C60	428	0.0502 -		SFAM01.1	PPB	6/27/2024 17:22	SMPL
P2845-18	MA4CJ1	8783	1.9147 -		SFAM01.1	PPB	6/27/2024 17:24	SMPL
P2845-19	MA4CJ2	4732	1.0107 -		SFAM01.1	PPB	6/27/2024 17:26	SMPL
CCV011	CCV011	22699	5.0202 -		SFAM01.1	PPB	6/27/2024 17:28	SMPL
CCB011	CCB011	-223	-0.0951 -		SFAM01.1	PPB	6/27/2024 17:31	SMPL
P2845-20	MA4CN3	209	0.0013 -		SFAM01.1	PPB	6/27/2024 17:33	SMPL
P2845-21	MA4C49	2352	0.4796 -		SFAM01.1	PPB	6/27/2024 17:35	SMPL
P2845-22	MA4C50	932	0.1627 -		SFAM01.1	PPB	6/27/2024 17:38	SMPL
P2845-03	MA4C64D	2234	0.4532 -		SFAM01.1	PPB	6/27/2024 17:43	SMPL
P2781-05DLX10	MA4D54	17263	3.8071 -		SFAM01.1	PPB	6/27/2024 17:45	SMPL
CCV012	CCV012	21853	4.8314 -		SFAM01.1	PPB	6/27/2024 17:53	SMPL
CCB012	CCB012	-18	-0.0493 -		SFAM01.1	PPB	6/27/2024 17:56	SMPL