

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB131442

Slope : 98.4

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	06/28/2024	09:55
2	CAL2	1	Water	NA	NA	20.3	7.00	06/28/2024	09:56
3	CAL3	1	Water	NA	NA	20.3	10.02	06/28/2024	09:59
4	ICV	1	Water	NA	NA	20.3	7.01	06/28/2024	10:00
5	CCV1	1	Water	NA	NA	20.2	2.01	06/28/2024	10:01
6	P3106-04	1	Solid	20.02	20	20.9	5.02	06/28/2024	10:10
7	P3106-04DUP	1	Solid	20.03	20	20.8	5.03	06/28/2024	10:11
8	CCV2	1	Water	NA	NA	20.3	12.02	06/28/2024	10:15

W3131442

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph p3106 solids

WorkList ID : 181481

Department : Wet-Chemistry

Date : 06-28-2024 09:36:52

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3106-04 B	FILTER-1	Solid	pH	Cool 4 deg C	PSEG03	M21	06/27/2024	9045D

Date/Time 06/28/24 09:40

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 06/28/24

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

141-00

[Signature]

[Signature]