

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB131444

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	06/28/2024	09:55
2	CAL2	1	Water	NA	NA	20.3	7.00	06/28/2024	09:56
3	CAL3	1	Water	NA	NA	20.3	10.02	06/28/2024	09:59
4	ICV	1	Water	NA	NA	20.3	7.01	06/28/2024	10:00
5	CCV1	1	Water	NA	NA	20.5	2.02	06/28/2024	11:10
6	P3106-06	1	Water	NA	NA	23.8	5.44	06/28/2024	11:15
7	P3106-06DUP	1	Water	NA	NA	23.9	5.45	06/28/2024	11:16
8	P3107-01	1	Water	NA	NA	24.0	7.51	06/28/2024	11:20
9	P3115-04	1	Water	NA	NA	20.7	7.24	06/28/2024	11:25
10	P3115-05	1	Water	NA	NA	20.3	7.16	06/28/2024	11:30
11	P3115-08	1	Water	NA	NA	20.2	6.78	06/28/2024	11:35
12	P3115-09	1	Water	NA	NA	20.7	4.87	06/28/2024	11:47
13	P3115-10	1	Water	NA	NA	20.5	6.08	06/28/2024	11:55
14	P3115-11	1	Water	NA	NA	20.3	7.89	06/28/2024	12:00
15	P3116-03	1	Water	NA	NA	20.1	9.07	06/28/2024	12:10
16	CCV2	1	Water	NA	NA	20.2	12.02	06/28/2024	12:15
17	P3120-01	1	Water	NA	NA	20.8	7.35	06/28/2024	12:22
18	CCV3	1	Water	NA	NA	20.2	2.02	06/28/2024	12:25

WORKLIST(Hardcopy Internal Chain)

VB131444

WorkList Name : ph water p2115 WorkList ID : 181480 Department : Wet-Chemistry Date : 06-28-2024 09:35:16

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3106-06	C 50370	Water	pH	Cool 4 deg C	PSEG03	M21	06/27/2024	9040C
P3107-01	A 50266	Water	pH	Cool 4 deg C	PSEG03	M21	06/27/2024	9040C
P3115-04	A D3566	Water	pH	Cool 4 deg C	PSEG03	M21	06/27/2024	9040C
P3115-05	A D3591-93	Water	pH	Cool 4 deg C	PSEG03	M21	06/27/2024	9040C
P3115-08	A D3586-88-90	Water	pH	Cool 4 deg C	PSEG03	M21	06/27/2024	9040C
P3115-09	A T3565	Water	pH	Cool 4 deg C	PSEG03	M21	06/27/2024	9040C
P3115-10	C T3558	Water	pH	Cool 4 deg C	PSEG03	M21	06/27/2024	9040C
P3115-11	A T3531	Water	pH	Cool 4 deg C	PSEG03	M21	06/27/2024	9040C
P3116-03	A D252	Water	pH	Cool 4 deg C	PSEG03	M21	06/27/2024	9040C
P3120-01	C WC-2040627	Water	pH	Cool 4 deg C	PARS02	M21	06/27/2024	9040C

Date/Time 06/28/24 10:00

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 06/28/24

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

14:00

Rm Sm

707 CeeC