

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB131447

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3085

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.7	06/28/2024	09:20
2	P3106-06		1	12	100.00	>212.0	>212.0	06/28/2024	09:50
3	P3106-06DUP		1	12	100.00	>212.0	>212.0	06/28/2024	10:20
4	P3107-01		1	14	100.00	>212.0	>212.0	06/28/2024	10:50
5	P3115-04		1	11	100.00	>212.0	>212.0	06/28/2024	11:20
6	P3115-05		1	14	100.00	>212.0	>212.0	06/28/2024	11:50
7	P3115-08		1	12	100.00	>212.0	>212.0	06/28/2024	12:20
8	P3115-09		1	14	100.00	>212.0	>212.0	06/28/2024	12:50
9	P3115-10		1	12	100.00	>212.0	>212.0	06/28/2024	13:20
10	P3115-11		1	13	100.00	>212.0	>212.0	06/28/2024	13:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB31447

WorkList Name : fp-6-28

WorkList ID : 181475

Department : Wet-Chemistry

Date : 06-28-2024 08:34:33

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3106-06	50370	Water	Flash Point	Cool 4 deg C	PSEG03	M21	06/27/2024	1010B
P3107-01	50266	Water	Flash Point	Cool 4 deg C	PSEG03	M21	06/27/2024	1010B
P3115-04	D3566	Water	Flash Point	Cool 4 deg C	PSEG03	M21	06/27/2024	1010B
P3115-05	D3591-93	Water	Flash Point	Cool 4 deg C	PSEG03	M21	06/27/2024	1010B
P3115-08	D3586-88-90	Water	Flash Point	Cool 4 deg C	PSEG03	M21	06/27/2024	1010B
P3115-09	T3565	Water	Flash Point	Cool 4 deg C	PSEG03	M21	06/27/2024	1010B
P3115-10	T3558	Water	Flash Point	Cool 4 deg C	PSEG03	M21	06/27/2024	1010B
P3115-11	T3531	Water	Flash Point	Cool 4 deg C	PSEG03	M21	06/27/2024	1010B

Date/Time 06/28/2024 09:10
 Raw Sample Received by: RM cws
 Raw Sample Relinquished by: JH(cws)

Date/Time 06/28/2024 14:36
 Raw Sample Received by: JH(cws)
 Raw Sample Relinquished by: RM(cws)