

LB 131448

SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

4473E

A= 0.0000e+000

B= 2.2337e-004 slope

C= -5.5284e-003 y-intercept

Rho= 0.9999281

Accept=Accepted

Max. Conc: —

10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	g/dl
0.00	0.000	0.009	0.009	63	0.000	63					
0.05	0.050					442	0				11
0.20	0.200	0.244	0.044	1118	0.0 %	1118					22
2.50	2.500	2.442	-0.058	10956	0.0 %	10956					11
5.00	5.000	4.954	-0.046	22204	0.0 %	22204					22
7.50	7.500	7.564	0.064	33887	0.0 %	33887					11
10.0	10.000	9.988	-0.012	44738	0.0 %	44738					0

LB131448

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	63	-		0 SFAM01.1	PPB	6/28/2024 15:58	Std
	0.2 S01	1118	-		0.2 SFAM01.1	PPB	6/28/2024 16:00	Std
	2.5 S02	10956	-		2.5 SFAM01.1	PPB	6/28/2024 16:02	Std
	5 S03	22204	-		5 SFAM01.1	PPB	6/28/2024 16:05	Std
	7.5 S04	33887	-		7.5 SFAM01.1	PPB	6/28/2024 16:07	Std
	10 S05	44738	-		10 SFAM01.1	PPB	6/28/2024 16:15	Std
ICV011	ICV011	17766	3.9629 -		SFAM01.1	PPB	6/28/2024 16:18	SMPL
ICB011	ICB011	17	-0.0017 -		SFAM01.1	PPB	6/28/2024 16:23	SMPL
CCV017	CCV017	22091	4.9289 -		SFAM01.1	PPB	6/28/2024 16:25	SMPL
CCB017	CCB017	30	0.0012 -		SFAM01.1	PPB	6/28/2024 16:30	SMPL
PB161836BL	PBW836	43	0.0041 -		SFAM01.1	PPB	6/28/2024 16:32	SMPL
P2409-02	MDL-WATER	296	0.0606 -		SFAM01.1	PPB	6/28/2024 16:37	SMPL
P3082-01	E08E0	52	0.0061 -		SFAM01.1	PPB	6/28/2024 16:39	SMPL
P3082-02	E08E0D	54	0.0065 -		SFAM01.1	PPB	6/28/2024 16:41	SMPL
P3082-03	E08E0S	4814	1.0698 -		SFAM01.1	PPB	6/28/2024 16:44	SMPL
P3083-01	E08E1	-13	-0.0084 -		SFAM01.1	PPB	6/28/2024 16:46	SMPL
P3083-02	E08E1D	46	0.0047 -		SFAM01.1	PPB	6/28/2024 16:48	SMPL
P3083-03	E08E1S	4664	1.0363 -		SFAM01.1	PPB	6/28/2024 16:51	SMPL
CCV018	CCV018	21573	4.8132 -		SFAM01.1	PPB	6/28/2024 16:53	SMPL
CCB018	CCB018	29	0.0009 -		SFAM01.1	PPB	6/28/2024 16:58	SMPL