

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB131450

Slope : 98.8

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	06/28/2024	16:00
2	CAL2	1	Water	NA	NA	20.3	7.01	06/28/2024	16:05
3	CAL3	1	Water	NA	NA	20.3	10.02	06/28/2024	16:10
4	ICV	1	Water	NA	NA	20.3	7.02	06/28/2024	16:15
5	CCV1	1	Water	NA	NA	20.3	2.01	06/28/2024	16:20
6	P3134-02	1	Solid	20.02	20	20.7	5.97	06/28/2024	16:25
7	P3134-02DUP	1	Solid	20.03	20	20.8	5.98	06/28/2024	16:26
8	P3135-06	1	Solid	20.05	20	20.7	8.40	06/28/2024	16:37
9	P3135-12	1	Solid	20.03	20	20.4	7.88	06/28/2024	16:45
10	CCV2	1	Water	NA	NA	20.4	12.02	06/28/2024	16:47

WORKLIST(Hardcopy Internal Chain)

VB131450

WorkList Name : corrosivity p3135 WorkList ID : 181535 Department : Wet-Chemistry Date : 06-28-2024 14:37:38

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3134-02	VNJ-238	Solid	Corrosivity	Cool 4 deg C	PSEG03	M11	06/28/2024	9045D
P3135-06	OK-628-COMP-01	Solid	Corrosivity	Cool 4 deg C	PSEG03	M21	06/28/2024	9045D
P3135-12	OK-628-COMP-02	Solid	Corrosivity	Cool 4 deg C	PSEG03	M21	06/28/2024	9045D

Date/Time 06/28/24 15:40

Raw Sample Received by: TD WPC

Raw Sample Relinquished by: ap sm

Date/Time 06/28/24 17:30

Raw Sample Received by: ap sm

Raw Sample Relinquished by: TD WPC