


LB131453

SFAM01.1 INSTRUMENT ID: CV1

Linear

 μ Abs.:

4356E



A= 0.0000e+000

B= 2.2748e-004 *slope*C= -3.7024e-002 *-100 intercept*

Rho= 0.9995394

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>%D</i>
0.00	0.000	-0.007	-0.007	130	0.000	130					<i>=</i>
0.05	0.050					442	0				<i>16</i>
0.20	0.200	0.231	0.031	1180	0.0 %	1180					<i>-6</i>
2.50	2.500	2.344	-0.156	10468	0.0 %	10468					<i>3</i>
5.00	5.000	5.146	0.146	22783	0.0 %	22783					<i>2</i>
7.50	7.500	7.613	0.113	33628	0.0 %	33628					<i>-1</i>
10.0	10.000	9.873	-0.127	43566	0.0 %	43566					

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0		130	-		0 SFAM01.1	PPB	6/28/2024 18:16	Std
0.2		1180	-		0.2 SFAM01.1	PPB	6/28/2024 18:19	Std
2.5		10468	-		2.5 SFAM01.1	PPB	6/28/2024 18:21	Std
5		22783	-		5 SFAM01.1	PPB	6/28/2024 18:26	Std
7.5		33628	-		7.5 SFAM01.1	PPB	6/28/2024 18:28	Std
10		43566	-		10 SFAM01.1	PPB	6/28/2024 18:31	Std
ICV012	ICV012	18987	4.2822 -		SFAM01.1	PPB	6/28/2024 18:33	SMPL
ICB012	ICB012	111	-0.0118 -		SFAM01.1	PPB	6/28/2024 18:38	SMPL
CCV019	CCV019	22828	5.1559 -		SFAM01.1	PPB	6/28/2024 18:41	SMPL
CCB019	CCB019	117	-0.0104 -		SFAM01.1	PPB	6/28/2024 18:45	SMPL
PB161839BL	PBS839	85	-0.0177 -		SFAM01.1	PPB	6/28/2024 18:50	SMPL
P2409-04	MDL-SOIL	337	0.0396 -		SFAM01.1	PPB	6/28/2024 18:53	SMPL
PB161840BL	PBS840	113	-0.0113 -		SFAM01.1	PPB	6/28/2024 19:00	SMPL
P2872-01	MA4C66	1981	0.4136 -		SFAM01.1	PPB	6/28/2024 19:02	SMPL
P2872-02	MA4C66D	2049	0.4291 -		SFAM01.1	PPB	6/28/2024 19:07	SMPL
P2872-03	MA4C66S	15237	3.4291 -		SFAM01.1	PPB	6/28/2024 19:10	SMPL
P2872-04	MA4C51	1188	0.2332 -		SFAM01.1	PPB	6/28/2024 19:12	SMPL
P2872-05	MA4CN6	1472	0.2978 -		SFAM01.1	PPB	6/28/2024 19:14	SMPL
P2872-06	MA4CN8	781	0.1406 -		SFAM01.1	PPB	6/28/2024 19:17	SMPL
P2872-07	MA4B58	14367	3.2312 -		SFAM01.1	PPB	6/28/2024 19:19	SMPL
P2872-08	MA4B59	1667	0.3422 -		SFAM01.1	PPB	6/28/2024 19:21	SMPL
P2872-09	MA4B73	12335	2.769 -		SFAM01.1	PPB	6/28/2024 19:23	SMPL
P2872-10	MA4CP8	9450	2.1127 -		SFAM01.1	PPB	6/28/2024 19:26	SMPL
CCV020	CCV020	22175	5.0074 -		SFAM01.1	PPB	6/28/2024 19:28	SMPL
CCB020	CCB020	63	-0.0227 -		SFAM01.1	PPB	6/28/2024 19:33	SMPL
P2872-11	MA4CP9	7018	1.5594 -		SFAM01.1	PPB	6/28/2024 19:35	SMPL
P2872-12	MA4CQ0	44643	10.1184 -		SFAM01.1	PPB	6/28/2024 19:37	SMPL
P2872-13	MA4CQ1	104082	23.6397 -		SFAM01.1	PPB	6/28/2024 19:40	SMPL
P2872-14	MA4CQ2	21697	4.8986 -		SFAM01.1	PPB	6/28/2024 19:42	SMPL
P2872-15	MA4B56	8101	1.8058 -		SFAM01.1	PPB	6/28/2024 19:44	SMPL
P2872-16	MA4CP0	16446	3.7041 -		SFAM01.1	PPB	6/28/2024 19:47	SMPL
P2872-17	MA4CP1	10663	2.3886 -		SFAM01.1	PPB	6/28/2024 19:49	SMPL
P2872-18	MA4CP2	9212	2.0585 -		SFAM01.1	PPB	6/28/2024 19:51	SMPL
P2872-19	MA4CP3	1562	0.3183 -		SFAM01.1	PPB	6/28/2024 19:53	SMPL
P2872-20	MA4CQ3	26761	6.0506 -		SFAM01.1	PPB	6/28/2024 19:56	SMPL
P2872-21	MA4CQ4	12236	2.7464 -		SFAM01.1	PPB	6/28/2024 19:58	SMPL
P2872-22	MA4CQ5	7719	1.7189 -		SFAM01.1	PPB	6/28/2024 20:00	SMPL
P2845-01DLX25	MA4D90	24082	5.4412 -		SFAM01.1	PPB	6/28/2024 20:07	SMPL
P2872-12DLX2		20907	4.7189 -		SFAM01.1	PPB	6/28/2024 20:12	SMPL
P2872-13DLX5		21217	4.7894 -		SFAM01.1	PPB	6/28/2024 20:14	SMPL
CCV021	CCV021	21869	4.9378 -		SFAM01.1	PPB	6/28/2024 20:16	SMPL
CCB021	CCB021	63	-0.0227 -		SFAM01.1	PPB	6/28/2024 20:21	SMPL