

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB131460

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3085

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	07/01/2024	14:35
2	P3110-07		1	4	7.00	44.6	44.3	07/01/2024	15:00
3	P3110-07DUP		1	4	7.00	44.6	44.3	07/01/2024	15:30
4	P3115-01		1	6	52.00	125.6	125.3	07/01/2024	16:00
5	P3116-02		1	6	53.00	127.4	127.1	07/01/2024	16:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

26131460.

WorkList Name : fp-s-6-28

WorkList ID : 181476

Department : Wet-Chemistry

Date : 06-28-2024 08:34:39

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3110-07 A	627	Solid	Flash Point	Cool 4 deg C	PSEG03	M11	06/27/2024	1010B
P3115-01 B	D3547	Solid	Flash Point	Cool 4 deg C	PSEG03	M21	06/27/2024	1010B
P3116-02 B	D249	Solid	Flash Point	Cool 4 deg C	PSEG03	M21	06/27/2024	1010B

Date/Time 07/01/2024 14:25
 Raw Sample Received by: RM
 Raw Sample Relinquished by: AD (woc)

Date/Time 07/01/2024 17:05
 Raw Sample Received by: AD (woc)
 Raw Sample Relinquished by: RM (woc)