

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB131468

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3085

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	07/01/2024	11:15
2	P3110-05		1	14	100.00	>212.0	>212.0	07/01/2024	11:45
3	P3110-05DUP		1	14	100.00	>212.0	>212.0	07/01/2024	12:15
4	P3110-06		1	12	100.00	>212.0	>212.0	07/01/2024	12:45
5	P3116-03		1	11	100.00	>212.0	>212.0	07/01/2024	13:15
6	P3120-01		1	14	100.00	>212.0	>212.0	07/01/2024	13:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6131468

WorkList Name : fp 7-1

WorkList ID : 181548

Department : Wet-Chemistry

Date : 07-01-2024 07:57:29

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3110-05 A	651	Water	Flash Point	Cool 4 deg C	PSEG03	M11	06/27/2024	1010B
P3110-06 A	652	Water	Flash Point	Cool 4 deg C	PSEG03	M11	06/27/2024	1010B
P3116-03 C	D252	Water	Flash Point	Cool 4 deg C	PSEG03	M21	06/27/2024	1010B
P3120-01 F	WC-2040627	Water	Flash Point	Cool 4 deg C	PARS02	M21	06/27/2024	1010B

Date/Time 07/01/2024 11.10
Raw Sample Received by: RM CWCJ
Raw Sample Relinquished by: JPL/CWCJ

Date/Time 07/01/2024 14.20
Raw Sample Received by: JPL/CWCJ
Raw Sample Relinquished by: JPL/CWCJ