

L3131470


SFAM01.1

INSTRUMENT ID: CV1

Linear

μ Abs.:

44260



A= 0.0000e+000

B= 2.2425e-004 slope

C= 6.6893e-002 y-intercept

Rho= 0.9997487

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	OK
0.00	0.000	0.064	0.064	-14	0.000	-14					
0.05	0.050					442	0				19
0.20	0.200	0.238	0.038	763	0.0 %	763					3
2.50	2.500	2.415	-0.085	10470	0.0 %	10470					2
5.00	5.000	4.877	-0.123	21449	0.0 %	21449					2
7.50	7.500	7.614	0.114	33656	0.0 %	33656					2
10.0	10.000	9.992	-0.008	44260	0.0 %	44260					0

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	-14	-		0 SFAM01.1	PPB	7/2/2024 8:56	Std
	0.2 S01	763	-		0.2 SFAM01.1	PPB	7/2/2024 8:58	Std
	2.5 S02	10470	-		2.5 SFAM01.1	PPB	7/2/2024 9:00	Std
	5 S03	21449	-		5 SFAM01.1	PPB	7/2/2024 9:02	Std
	7.5 S04	33656	-		7.5 SFAM01.1	PPB	7/2/2024 9:05	Std
	10 S05	44260	-		10 SFAM01.1	PPB	7/2/2024 9:10	Std
ICV014	ICV014	17957	4.0938	-	SFAM01.1	PPB	7/2/2024 9:13	SMPL
ICB014	ICB014	-177	0.0272	-	SFAM01.1	PPB	7/2/2024 9:15	SMPL
CCV026	CCV026	21876	4.9726	-	SFAM01.1	PPB	7/2/2024 9:17	SMPL
CCB026	CCB026	-164	0.0301	-	SFAM01.1	PPB	7/2/2024 9:19	SMPL
PB161879BL	PBS879	-11	0.0644	-	SFAM01.1	PPB	7/2/2024 9:22	SMPL
P2910-01	MA4B75	5015	1.1915	-	SFAM01.1	PPB	7/2/2024 9:24	SMPL
P2910-02	MA4B76	1462	0.3947	-	SFAM01.1	PPB	7/2/2024 9:26	SMPL
P2910-03	MA4B78	2124	0.5432	-	SFAM01.1	PPB	7/2/2024 9:28	SMPL
P2910-04	MA4B81	4941	1.1749	-	SFAM01.1	PPB	7/2/2024 9:31	SMPL
P2910-05	MA4B84	8677	2.0127	-	SFAM01.1	PPB	7/2/2024 9:33	SMPL
P2910-06	MA4B84D	8136	1.8914	-	SFAM01.1	PPB	7/2/2024 9:35	SMPL
P2910-07	MA4B84S	19740	4.4936	-	SFAM01.1	PPB	7/2/2024 9:38	SMPL
P2910-08	MA4B85	682	0.2198	-	SFAM01.1	PPB	7/2/2024 9:40	SMPL
P2910-09	MA4B41	3341	0.8161	-	SFAM01.1	PPB	7/2/2024 9:42	SMPL
P2910-10	MA4B42	5749	1.3561	-	SFAM01.1	PPB	7/2/2024 9:44	SMPL
P2910-11	MA4B43	2283	0.5789	-	SFAM01.1	PPB	7/2/2024 9:47	SMPL
P2910-12	MA4CQ6	19266	4.3873	-	SFAM01.1	PPB	7/2/2024 9:49	SMPL
P2910-13	MA4CQ7	19756	4.4972	-	SFAM01.1	PPB	7/2/2024 9:51	SMPL
P2910-14	MA4CQ8	19227	4.3786	-	SFAM01.1	PPB	7/2/2024 9:54	SMPL
P2910-15	MA4CQ9	16814	3.8375	-	SFAM01.1	PPB	7/2/2024 9:56	SMPL
P2910-16	MA4CR0	14431	3.3031	-	SFAM01.1	PPB	7/2/2024 9:58	SMPL
P2910-17	MA4BQ2	22452	5.1018	-	SFAM01.1	PPB	7/2/2024 10:00	SMPL
P2910-18	MA4CR5	10472	2.4153	-	SFAM01.1	PPB	7/2/2024 10:03	SMPL
P2910-19	MA4CR6	12382	2.8436	-	SFAM01.1	PPB	7/2/2024 10:05	SMPL
CCV027	CCV027	21678	4.9282	-	SFAM01.1	PPB	7/2/2024 10:07	SMPL
CCB027	CCB027	-168	0.0292	-	SFAM01.1	PPB	7/2/2024 10:10	SMPL
P2910-20	MA4CR7	46892	10.5825	-	SFAM01.1	PPB	7/2/2024 10:12	SMPL
P2910-21	MA4CR8	12689	2.9124	-	SFAM01.1	PPB	7/2/2024 10:14	SMPL
P2910-22	MA4CR9	26485	6.0062	-	SFAM01.1	PPB	7/2/2024 10:16	SMPL
P2910-23	MA4D67	326850	HIGH	-	SFAM01.1	PPB	7/2/2024 10:19	
P2910-20DLX2	MA4CR7	23069	5.2402	-	SFAM01.1	PPB	7/2/2024 10:26	SMPL
P2910-23DLX25	MA4D67	21715	4.9365	-	SFAM01.1	PPB	7/2/2024 10:33	SMPL
CCV028	CCV028	20879	4.749	-	SFAM01.1	PPB	7/2/2024 10:38	SMPL
CCB028	CCB028	-104	0.0436	-	SFAM01.1	PPB	7/2/2024 10:40	SMPL