

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 7/3/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 13:05
In Date: 07/02/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:00
Out Date: 07/03/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLIDS-OVEN

QC:LB131474

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
P3089-01	MA4CB7	1	1.18	8.66	9.84	7.9	77.6	
P3089-02	MA4CB7D	2	1.18	8.66	9.84	7.9	77.6	
P3089-03	MA4CB7S	3	1.18	8.66	9.84	7.9	77.6	
P3089-04	MA4BC6	4	1.12	8.57	9.69	5.42	50.2	
P3089-05	MA4BP1	5	1.18	8.45	9.63	4.5	39.3	
P3089-06	MA4BQ8	6	1.14	8.85	9.99	5.56	49.9	
P3089-07	MA4BR0	7	1.18	8.67	9.85	6.96	66.7	
P3089-08	MA4BR1	8	1.12	8.76	9.88	4.9	43.2	
P3089-09	MA4BR2	9	1.15	8.83	9.98	5.91	53.9	
P3089-10	MA4BR3	10	1.15	8.40	9.55	6.67	65.7	
P3089-11	MA4BR4	11	1.18	8.43	9.61	6.73	65.8	
P3089-12	MA4BR5	12	1.16	8.40	9.56	5.36	50.0	
P3089-13	MA4BR6	13	1.15	8.83	9.98	6.92	65.3	
P3089-14	MA4BR7	14	1.15	8.82	9.97	6.82	64.3	
P3089-15	MA4BR8	15	1.14	8.81	9.95	6.66	62.7	
P3089-16	MA4D02	16	1.18	8.44	9.62	6.85	67.2	
P3089-17	MA4D11	17	1.14	8.83	9.97	8.55	83.9	
P3089-18	MA4BX1	18	1.18	8.79	9.97	5.47	48.8	
P3089-19	MA4BY0	19	1.18	8.77	9.95	5.18	45.6	
P3089-20	MA4D38	20	1.11	8.57	9.68	9.28	95.3	
P3089-21	MA4D51	21	1.16	8.49	9.65	8.96	91.9	
P3089-22	MA4D52	22	1.18	8.44	9.62	6.38	61.6	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$