

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB131480

Slope : 98.5

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	07/02/2024	16:15
2	CAL2	1	Water	NA	NA	20.3	7.00	07/02/2024	16:16
3	CAL3	1	Water	NA	NA	20.3	10.02	07/02/2024	16:20
4	ICV	1	Water	NA	NA	20.3	7.01	07/02/2024	16:25
5	CCV1	1	Water	NA	NA	20.3	2.01	07/02/2024	16:30
6	P3152-01	1	Solid	20.02	20	21.3	7.26	07/02/2024	16:38
7	P3152-01DUP	1	Solid	20.03	20	21.4	7.27	07/02/2024	16:39
8	P3152-05	1	Solid	20.05	20	21.6	8.13	07/02/2024	16:45
9	CCV2	1	Water	NA	NA	20.3	12.02	07/02/2024	16:48

WORKLIST(Hardcopy Internal Chain)

2131480

WorkList Name : corrosivity p3152 WorkList ID : 181593 Department : Wet-Chemistry Date : 07-02-2024 15:36:27

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3152-01	441-443	Solid	Corrosivity	Cool 4 deg C	PSEG03	D11	07/02/2024	9045D
P3152-05	444-446	Solid	Corrosivity	Cool 4 deg C	PSEG03	D11	07/02/2024	9045D

Date/Time 07/02/24 15:40
Raw Sample Received by: 70609C
Raw Sample Relinquished by: CP SR

Date/Time 07/02/24 17:30
Raw Sample Received by: CP SR
Raw Sample Relinquished by: 70609C