

Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Niha

Parameter: Turbidity

SUPERVISOR REVIEW BY: Iwona

Run Number: LB131487

Reagent/Standard	Lot/Log #
Turbidity Calibration std, 0NTU	WP108615
Turbidity Calibration std, 20NTU	WP108616
Turbidity Calibration std, 40NTU	WP108617
Turbidity Calibration std, 1NTU	WP108618
Turbidity Calibration std, 5NTU	WP108619
Turbidity Calibration - CCV std, 10 NTU	WP108620
10 NTU Standard 500 ml	W2971

Intercept: -0.0899

Slope: 0.9798

Regression: 0.9999

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	0.149	0.24		07/03/2024	08:40
2	CAL2	1	1	1.120	1.24	23.5	07/03/2024	08:44
3	CAL3	5	1	4.480	4.66	-6.7	07/03/2024	08:48
4	CAL4	10	1	9.480	9.77	-2.3	07/03/2024	08:52
5	CAL5	20	1	19.500	19.99	-0	07/03/2024	08:56
6	CAL6	40	1	39.200	40.10	0.3	07/03/2024	09:00



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Niha

Parameter: Turbidity

SUPERVISOR REVIEW BY: Iwona

Run Number: LB131487

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.820	10.114	07/03/2024	09:04
2	ICB		1	0.160	0.255	07/03/2024	09:08
3	CCV1	10	1	9.800	10.094	07/03/2024	09:12
4	CCB1		1	0.160	0.255	07/03/2024	09:16
5	LB131487BL		1	0.150	0.245	07/03/2024	09:20
6	P3153-01		1	0.689	0.795	07/03/2024	09:24
7	P3153-01DUP		1	0.692	0.798	07/03/2024	09:28
8	P3153-02		1	1.190	1.306	07/03/2024	09:32
9	P3153-03		1	1.710	1.837	07/03/2024	09:36
10	P3153-04		1	0.820	0.929	07/03/2024	09:40
11	P3153-05		1	0.636	0.741	07/03/2024	09:44
12	P3153-06		1	0.758	0.865	07/03/2024	09:48
13	P3153-07		1	0.950	1.061	07/03/2024	09:52
14	P3153-08		1	0.706	0.812	07/03/2024	09:56
15	CCV2	10	1	9.700	9.992	07/03/2024	10:00
16	CCB2		1	0.160	0.255	07/03/2024	10:04

WORKLIST(Hardcopy Internal Chain)

LB131487

WorkList Name : TURBIDITY-07022024

WorkList ID : 181595

Department : Wet-Chemistry

Date : 07-02-2024 15:45:36

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3153-01	WO70224RB01	Water	Turbidity	Cool 4 deg C	CLOU03	D11	07/02/2024	SM2130 B
P3153-02	WO70224RB02	Water	Turbidity	Cool 4 deg C	CLOU03	D11	07/02/2024	SM2130 B
P3153-03	WO70224RB03	Water	Turbidity	Cool 4 deg C	CLOU03	D11	07/02/2024	SM2130 B
P3153-04	WO70224RB04	Water	Turbidity	Cool 4 deg C	CLOU03	D11	07/02/2024	SM2130 B
P3153-05	WO70224RB05	Water	Turbidity	Cool 4 deg C	CLOU03	D11	07/02/2024	SM2130 B
P3153-06	WO70224RB06	Water	Turbidity	Cool 4 deg C	CLOU03	D11	07/02/2024	SM2130 B
P3153-07	WO70224RB07	Water	Turbidity	Cool 4 deg C	CLOU03	D11	07/02/2024	SM2130 B
P3153-08	WO70224RB08	Water	Turbidity	Cool 4 deg C	CLOU03	D11	07/02/2024	SM2130 B

Date/Time 07-03-2024 . 08:30

Raw Sample Received by: NF (WC)

Raw Sample Relinquished by: [Signature]

Date/Time 07-03-2024 . 10:10

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: NF (WC)