


LB131489

7471B

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 1.8615e-004

C= 1.0534e-002

Rho= 0.9999895

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.019	0.019	46	0.000	46					
0.2	0.200	0.202	0.002	1028	0.0 %	1028					
2.5	2.500	2.491	-0.009	13323	0.0 %	13323					
5.0	5.000	4.968	-0.032	26630	0.0 %	26630					
7.5	7.500	7.510	0.010	40286	0.0 %	40286					
10.0	10.000	10.011	0.011	53723	0.0 %	53723					

100.000

LB131489 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 03 Jul 2024 11:43:05

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	46	-	0.0000	7471B	PPB	03 Jul 2024 12:02:41	S	Std
0.2 - 1	50.2	1028	-	0.2000	7471B	PPB	03 Jul 2024 12:04:58	S	Std
2.5 - 1	52.5	13323	-	2.5000	7471B	PPB	03 Jul 2024 12:07:16	S	Std
5.0 - 1	55	26630	-	5.0000	7471B	PPB	03 Jul 2024 12:09:33	S	Std
7.5 - 1	57.5	40286	-	7.5000	7471B	PPB	03 Jul 2024 12:11:52	S	Std
10.0 - 1	590	53723	-	10.0000	7471B	PPB	03 Jul 2024 12:14:12	S	Std
ICV16 - 1	ICV16	20857	3.8930	-	7471B	PPB	03 Jul 2024 12:17:07	U	SMPL
ICB16 - 1	ICB16	-336	-0.0520	-	7471B	PPB	03 Jul 2024 12:19:23	U	SMPL
CCV35 - 1	CCV35	27121	5.0591	-	7471B	PPB	03 Jul 2024 12:21:43	U	SMPL
CCB35 - 1	CCB35	-360	-0.0565	-	7471B	PPB	03 Jul 2024 12:24:01	U	SMPL
CRA - 1	CRA	979	0.1928	-	7471B	PPB	03 Jul 2024 12:26:20	U	SMPL
HighStd - 1	HighStd	54101	10.0814	-	7471B	PPB	03 Jul 2024 12:28:42	U	SMPL
ChkStd - 1	ChkStd	38381	7.1551	-	7471B	PPB	03 Jul 2024 12:30:58	U	SMPL
PB161900BL - 1	PB161900BL	-425	-0.0686	-	7471B	PPB	03 Jul 2024 12:33:25	U	SMPL
PB161900BS - 1	PB161900BS	19662	3.6706	-	7471B	PPB	03 Jul 2024 12:39:25	U	SMPL
P3141-01 - 1	B-140-SB00	15159	2.8324	-	7471B	PPB	03 Jul 2024 12:41:42	U	SMPL
P3141-02 - 1	B-140-SB01	6364	1.1952	-	7471B	PPB	03 Jul 2024 12:44:02	U	SMPL
P3141-03 - 1	B-140-SB02	2732	0.5191	-	7471B	PPB	03 Jul 2024 12:46:24	U	SMPL
P3141-04 - 1	B-129-SB01	690	0.1390	-	7471B	PPB	03 Jul 2024 12:48:45	U	SMPL
P3141-05 - 1	B-129-SB02	2451	0.4668	-	7471B	PPB	03 Jul 2024 12:51:04	U	SMPL
CCV36 - 1	CCV36	26086	4.8664	-	7471B	PPB	03 Jul 2024 12:53:19	U	SMPL
CCB36 - 1	CCB36	29	0.0159	-	7471B	PPB	03 Jul 2024 12:58:53	U	SMPL
P3142-01 - 1	VNJ-231	1124	0.2198	-	7471B	PPB	03 Jul 2024 13:01:08	U	SMPL
P3142-03 - 1	327	11790	2.2052	-	7471B	PPB	03 Jul 2024 13:03:24	U	SMPL
P3148-01 - 1	EO-02-070224	2914	0.5530	-	7471B	PPB	03 Jul 2024 13:05:39	U	SMPL
P3149-01 - 1	OK-01-070224	1249	0.2430	-	7471B	PPB	03 Jul 2024 13:07:57	U	SMPL
P3149-01DUP - 1	OK-01-070224DUP	1157	0.2259	-	7471B	PPB	03 Jul 2024 13:10:15	U	SMPL
P3149-01MS - 1	OK-01-070224MS	21436	4.0008	-	7471B	PPB	03 Jul 2024 13:12:31	U	SMPL
P3149-01MSD - 1	OK-01-070224MSD	21502	4.0131	-	7471B	PPB	03 Jul 2024 13:14:48	U	SMPL
P3149-01LX5 - 1		272	0.0612	-	7471B	PPB	03 Jul 2024 13:19:39	U	SMPL
P3149-01A - 1		21747	4.0587	-	7471B	PPB	03 Jul 2024 13:21:56	U	SMPL
CCV37 - 1	CCV37	26902	5.0183	-	7471B	PPB	03 Jul 2024 13:24:11	U	SMPL
CCB37 - 1	CCB37	-413	-0.0663	-	7471B	PPB	03 Jul 2024 13:26:41	U	SMPL