


LB131506

SFAM01.1

INSTRUMENT ID: EV1



Linear

A= 0.0000e+000

B= 2.0360e-004 slope

C= -1.5823e-002 y-intercept

Rho= 0.9997646

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	9-b
0.00	0.000	0.052	0.052	331	0.000	331					
0.05	0.050					442	0				
0.20	0.200	0.225	0.025	1181	0.0 %	1181					
2.50	2.500	2.481	-0.019	12262	0.0 %	12262					
5.00	5.000	4.839	-0.161	23844	0.0 %	23844					
7.50	7.500	7.593	0.093	37373	0.0 %	37373					
10.0	10.000	10.011	0.011	49250	0.0 %	49250					

9-b
= 13
- 1
- 3
- 1
0

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	331	-		0 SFAM01.1	PPB	7/3/2024 16:26	Std
	0.2 S01	1181	-		0.2 SFAM01.1	PPB	7/3/2024 16:29	Std
	2.5 S02	12262	-		2.5 SFAM01.1	PPB	7/3/2024 16:31	Std
	5 S03	23844	-		5 SFAM01.1	PPB	7/3/2024 16:33	Std
	7.5 S04	37373	-		7.5 SFAM01.1	PPB	7/3/2024 16:36	Std
	10 S05	49250	-		10 SFAM01.1	PPB	7/3/2024 16:38	Std
ICV017	ICV017	19606	3.9759 -		SFAM01.1	PPB	7/3/2024 16:41	SMPL
ICB017	ICB017	193	0.0235 -		SFAM01.1	PPB	7/3/2024 16:43	SMPL
CCV038	CCV038	24144	4.8998 -		SFAM01.1	PPB	7/3/2024 16:45	SMPL
CCB038	CCB038	169	0.0186 -		SFAM01.1	PPB	7/3/2024 16:48	SMPL
PB161912BL	PBS912	15	-0.0128 -		SFAM01.1	PPB	7/3/2024 16:50	SMPL
P2935-01	MA4B61	1221	0.2328 -		SFAM01.1	PPB	7/3/2024 16:52	SMPL
P2935-02	MA4B62	785	0.144 -		SFAM01.1	PPB	7/3/2024 16:54	SMPL
P2935-03	MA4B72	4521	0.9046 -		SFAM01.1	PPB	7/3/2024 16:57	SMPL
P2935-04	MA4B72D	4014	0.8014 -		SFAM01.1	PPB	7/3/2024 17:02	SMPL
P2935-05	MA4B72S	15721	3.1849 -		SFAM01.1	PPB	7/3/2024 17:04	SMPL
P2935-06	MA4CP4	10905	2.2044 -		SFAM01.1	PPB	7/3/2024 17:06	SMPL
P2935-07	MA4CP5	19661	3.9871 -		SFAM01.1	PPB	7/3/2024 17:09	SMPL
P2935-08	MA4CP6	8442	1.7029 -		SFAM01.1	PPB	7/3/2024 17:11	SMPL
P2935-09	MA4CP7	9505	1.9194 -		SFAM01.1	PPB	7/3/2024 17:13	SMPL
P2935-10	MA4B57	14564	2.9493 -		SFAM01.1	PPB	7/3/2024 17:15	SMPL
P2935-11	MA4D91	348786	HIGH -		SFAM01.1	PPB	7/3/2024 17:18	
P2935-12	MA4B44	1858	0.3625 -		SFAM01.1	PPB	7/3/2024 17:20	SMPL
P2935-13	MA4B45	18514	3.7535 -		SFAM01.1	PPB	7/3/2024 17:22	SMPL
P2935-14	MA4B46	20690	4.1966 -		SFAM01.1	PPB	7/3/2024 17:24	SMPL
P2935-15	MA4B47	4048	0.8083 -		SFAM01.1	PPB	7/3/2024 17:26	SMPL
P2935-16	MA4B48	1386	0.2664 -		SFAM01.1	PPB	7/3/2024 17:29	SMPL
P2935-17	MA4CR1	27355	5.5535 -		SFAM01.1	PPB	7/3/2024 17:31	SMPL
P2935-18	MA4CR3	29132	5.9153 -		SFAM01.1	PPB	7/3/2024 17:33	SMPL
P2935-19	MA4CS7	16145	3.2712 -		SFAM01.1	PPB	7/3/2024 17:36	SMPL
CCV039	CCV039	23658	4.8008 -		SFAM01.1	PPB	7/3/2024 17:38	SMPL
CCB039	CCB039	97	0.0039 -		SFAM01.1	PPB	7/3/2024 17:40	SMPL
P2935-20	MA4CS8	15094	3.0573 -		SFAM01.1	PPB	7/3/2024 17:43	SMPL
P2935-21	MA4CS9	17098	3.4653 -		SFAM01.1	PPB	7/3/2024 17:45	SMPL
P2935-22	MA4B49	13314	2.6949 -		SFAM01.1	PPB	7/3/2024 17:47	SMPL
P2935-23	MA4B50	3290	0.654 -		SFAM01.1	PPB	7/3/2024 17:49	SMPL
PB161913BL	PBS913	-46	-0.0252 -		SFAM01.1	PPB	7/3/2024 17:52	SMPL
P2965-01	MA4B51	3630	0.7232 -		SFAM01.1	PPB	7/3/2024 17:54	SMPL
P2965-02	MA4B52	8915	1.7992 -		SFAM01.1	PPB	7/3/2024 17:56	SMPL
P2965-03	MA4B53	11338	2.2925 -		SFAM01.1	PPB	7/3/2024 17:59	SMPL
P2965-04	MA4B54	4850	0.9716 -		SFAM01.1	PPB	7/3/2024 18:01	SMPL
P2965-05	MA4CS2	35722	7.257 -		SFAM01.1	PPB	7/3/2024 18:03	SMPL
P2965-06	MA4CS3	35936	7.3006 -		SFAM01.1	PPB	7/3/2024 18:05	SMPL
P2965-07	MA4CS3D	35916	7.2965 -		SFAM01.1	PPB	7/3/2024 18:10	SMPL
P2965-08	MA4CS3S	46168	9.3838 -		SFAM01.1	PPB	7/3/2024 18:15	SMPL

INSTRUMENT ID : CV1

P2965-09	MA4CS5	51134	10.3948 -	SFAM01.1	PPB	7/3/2024 18:18	SMPL
P2965-10	MA4CS6	14106	2.8561 -	SFAM01.1	PPB	7/3/2024 18:20	SMPL
P2965-11	MA4CT0	23927	4.8556 -	SFAM01.1	PPB	7/3/2024 18:22	SMPL
P2965-12	MA4CT3	16714	3.3871 -	SFAM01.1	PPB	7/3/2024 18:25	SMPL
P2965-13	MA4CT4	8602	1.7355 -	SFAM01.1	PPB	7/3/2024 18:27	SMPL
P2965-14	MA4CT5	13517	2.7362 -	SFAM01.1	PPB	7/3/2024 18:29	SMPL
P2965-15	MA4CT6	18367	3.7236 -	SFAM01.1	PPB	7/3/2024 18:31	SMPL
CCV040	CCV040	23444	4.7573 -	SFAM01.1	PPB	7/3/2024 18:34	SMPL
CCB040	CCB040	156	0.0159 -	SFAM01.1	PPB	7/3/2024 18:36	SMPL
P2965-16	MA4CT7	12582	2.5458 -	SFAM01.1	PPB	7/3/2024 18:38	SMPL
P2965-18	MA4D68	237159	48.2687 -	SFAM01.1	PPB	7/3/2024 18:41	SMPL
P2965-19	MA4CF2	4666	0.9342 -	SFAM01.1	PPB	7/3/2024 18:43	SMPL
P2965-20	MA4CF4	2130	0.4178 -	SFAM01.1	PPB	7/3/2024 18:45	SMPL
P2965-21	MA4CF5	484	0.0827 -	SFAM01.1	PPB	7/3/2024 18:48	SMPL
P2965-22	MA4CF6	10480	2.1179 -	SFAM01.1	PPB	7/3/2024 18:50	SMPL
P2965-23	MA4CF7	1796	0.3498 -	SFAM01.1	PPB	7/3/2024 18:52	SMPL
P2935-11DLX25	MA4D91	25399	5.1553 -	SFAM01.1	PPB	7/3/2024 19:11	SMPL
P2965-18DLX10	MA4D68	26159	5.31 -	SFAM01.1	PPB	7/3/2024 19:20	SMPL
CCV041	CCV041	21495	4.3605 -	SFAM01.1	PPB	7/3/2024 19:27	SMPL
CCB041	CCB041	170	0.0188 -	SFAM01.1	PPB	7/3/2024 19:29	SMPL
P2965-09DLX2		22458	4.5565 -	SFAM01.1	PPB	7/3/2024 20:07	SMPL
CCV042	CCV042	22058	4.4751 -	SFAM01.1	PPB	7/3/2024 20:09	SMPL
CCB042	CCB042	144	0.0135 -	SFAM01.1	PPB	7/3/2024 20:12	SMPL