

LB131517


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

37933



A= 0.0000e+000

B= 2.5946e-004

C= 2.2504e-002

Rho= 0.9992483

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/1/2
0.0	0.000	0.051	0.051	111	0.000	111	0				
0.2	0.200	0.213	0.013	735	0.0 %	735					
2.5	2.500	2.268	-0.232	8656	0.0 %	8656					
5.0	5.000	5.184	0.184	19895	0.0 %	19895					
7.5	7.500	7.618	0.118	29275	0.0 %	29275					
10.0	10.000	9.865	-0.135	37933	0.0 %	37933					

LB131517 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 05 Jul 2024 13:03:48

Sample ID	Extended ID	μ Abs	Conc	Stnd Conc	Method	Units	Date	Type	Type
0.0 - 1	50	111	-	0.0000	7470A	PPB	05 Jul 2024 13:41:13	S	Std
0.2 - 1	50-2	735	-	0.2000	7470A	PPB	05 Jul 2024 13:43:30	S	Std
2.5 - 1	52-5	8656	-	2.5000	7470A	PPB	05 Jul 2024 13:45:48	S	Std
5.0 - 1	55	18895	-	5.0000	7470A	PPB	05 Jul 2024 13:48:05	S	Std
7.5 - 1	55-5	29275	-	7.5000	7470A	PPB	05 Jul 2024 13:50:21	S	Std
10.0 - 1	516	37933	-	10.0000	7470A	PPB	05 Jul 2024 13:52:39	S	Std
ICV18 - 1	ICV18	16675	4.3490	-	7470A	PPB	05 Jul 2024 13:58:06	U	SMPL
ICB18 - 1	ICB18	-106	-0.0050	-	7470A	PPB	05 Jul 2024 14:00:21	U	SMPL
CCV43 - 1	CCV43	19107	4.9800	-	7470A	PPB	05 Jul 2024 14:02:38	U	SMPL
CCB43 - 1	CCB43	-55	0.0082	-	7470A	PPB	05 Jul 2024 14:04:53	U	SMPL
CRA - 1	CRA	635	0.1873	-	7470A	PPB	05 Jul 2024 14:07:10	U	SMPL
HighStd - 1	HighStd	37730	9.8119	-	7470A	PPB	05 Jul 2024 14:09:25	U	SMPL
ChkStd - 1	ChkStd	26898	7.0014	-	7470A	PPB	05 Jul 2024 14:11:40	U	SMPL
PB161918BL - 1	PBW	-373	-0.0743	-	7470A	PPB	05 Jul 2024 14:13:58	U	SMPL
PB161918BS - 1	LCSW	17195	4.4839	-	7470A	PPB	05 Jul 2024 14:16:17	U	SMPL
P3142-02 - 1	VNJ-231	-252	-0.0429	-	7470A	PPB	05 Jul 2024 14:18:34	U	SMPL
P3142-04 - 1	327	-64	0.0059	-	7470A	PPB	05 Jul 2024 14:20:52	U	SMPL
P3152-01 - 1	441-443	3993	1.0585	-	7470A	PPB	05 Jul 2024 14:23:10	U	SMPL
P3152-05 - 1	444-446	2122	0.5731	-	7470A	PPB	05 Jul 2024 14:25:27	U	SMPL
P3152-05DUP - 1	444-446DUP	2455	0.6595	-	7470A	PPB	05 Jul 2024 14:27:43	U	SMPL
CCV44 - 1	CCV44	19955	5.2000	-	7470A	PPB	05 Jul 2024 14:30:00	U	SMPL
CCB44 - 1	CCB44	-66	0.0054	-	7470A	PPB	05 Jul 2024 14:32:16	U	SMPL
P3152-05MS - 1	444-446MS	16648	4.3420	-	7470A	PPB	05 Jul 2024 14:34:33	U	SMPL
P3152-05MSD - 1	444-446MSD	19691	5.1315	-	7470A	PPB	05 Jul 2024 14:36:49	U	SMPL
P3157-02 - 1	PALISADES-POLE-YARD-DG-PAINT	-216	-0.0335	-	7470A	PPB	05 Jul 2024 14:39:06	U	SMPL
P3157-05 - 1	PALISADES-POLE-YARD-LG-PAINT	1291	0.3575	-	7470A	PPB	05 Jul 2024 14:41:24	U	SMPL
P3157-08 - 1	PALISADES-POLE-YARD-Y-PAINT	358	0.1154	-	7470A	PPB	05 Jul 2024 14:43:40	U	SMPL
P3157-11 - 1	PALISADES-POLE-YARD-W-PAINT	173	0.0674	-	7470A	PPB	05 Jul 2024 14:45:56	U	SMPL
PB161878TB - 1	PB161878TB	-13	0.0191	-	7470A	PPB	05 Jul 2024 14:48:13	U	SMPL
PB161911TB - 1	PB161911TB	2	0.0230	-	7470A	PPB	05 Jul 2024 14:50:30	U	SMPL
P3152-05LX5 - 1		334	0.1092	-	7470A	PPB	05 Jul 2024 14:57:58	U	SMPL
P3152-05A - 1		15336	4.0016	-	7470A	PPB	05 Jul 2024 15:00:13	U	SMPL
CCV45 - 1	CCV45	19942	5.1966	-	7470A	PPB	05 Jul 2024 15:02:29	U	SMPL
CCB45 - 1	CCB45	-212	-0.0325	-	7470A	PPB	05 Jul 2024 15:04:48	U	SMPL