


LB131526

7471B

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 1.7501e-004

C= 1.4883e-002

Rho= 0.9998194

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.031	0.031	91	0.000	91					
0.2	0.200	0.214	0.014	1140	0.0 %	1140					
2.5	2.500	2.502	0.002	14213	0.0 %	14213					
5.0	5.000	4.863	-0.137	27704	0.0 %	27704					
7.5	7.500	7.598	0.098	43329	0.0 %	43329					
10.0	10.000	9.991	-0.009	57001	0.0 %	57001					

170310

LB131526 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 08 Jul 2024 08:57:01

Sample ID	Extended ID	µ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	91	-	0.0000	7471B	PPB	08 Jul 2024 09:17:03	S	Std
0.2 - 1	50-2	1140	-	0.2000	7471B	PPB	08 Jul 2024 09:19:20	S	Std
2.5 - 1	52.5	14213	-	2.5000	7471B	PPB	08 Jul 2024 09:21:37	S	Std
5.0 - 1	55	27704	-	5.0000	7471B	PPB	08 Jul 2024 09:23:54	S	Std
7.5 - 1	57.5	43329	-	7.5000	7471B	PPB	08 Jul 2024 09:26:13	S	Std
10.0 - 1	59	57001	-	10.0000	7471B	PPB	08 Jul 2024 09:31:16	S	Std
ICV19 - 1	ICV19	23025	4.0446	-	7471B	PPB	08 Jul 2024 09:34:02	U	SMPL
ICB19 - 1	ICB19	-180	-0.0166	-	7471B	PPB	08 Jul 2024 09:36:18	U	SMPL
CCV46 - 1	CCV46	26993	4.7390	-	7471B	PPB	08 Jul 2024 09:38:36	U	SMPL
CCB46 - 1	CCB46	-215	-0.0227	-	7471B	PPB	08 Jul 2024 09:40:51	U	SMPL
CRA - 1	CRA	1106	0.2084	-	7471B	PPB	08 Jul 2024 09:43:09	U	SMPL
HighStd - 1	HighStd	56772	9.9508	-	7471B	PPB	08 Jul 2024 09:45:24	U	SMPL
ChkStd - 1	ChkStd	40798	7.1551	-	7471B	PPB	08 Jul 2024 09:47:40	U	SMPL
PB161941BL - 1	PBS	-323	-0.0416	-	7471B	PPB	08 Jul 2024 09:50:04	U	SMPL
PB161941BS - 1	LCSS	23594	4.1442	-	7471B	PPB	08 Jul 2024 09:52:24	U	SMPL
P3076-02 - 1	7-N-2	9067	1.6017	-	7471B	PPB	08 Jul 2024 09:54:41	U	SMPL
P3076-02DUP - 1	7-N-2DUP	7185	1.2724	-	7471B	PPB	08 Jul 2024 09:57:00	U	SMPL
P3076-02MS - 1	7-N-2MS	31126	5.4624	-	7471B	PPB	08 Jul 2024 09:59:20	U	SMPL
P3076-02MSD - 1	7-N-2MSD	28811	5.0572	-	7471B	PPB	08 Jul 2024 10:01:39	U	SMPL
P3076-03 - 1	7-N-3	5459	0.9703	-	7471B	PPB	08 Jul 2024 10:03:58	U	SMPL
CCV47 - 1	CCV47	26089	4.5808	-	7471B	PPB	08 Jul 2024 10:06:18	U	SMPL
CCB47 - 1	CCB47	-168	-0.0145	-	7471B	PPB	08 Jul 2024 10:08:35	U	SMPL
P3076-05 - 1	7-W-2	6993	1.2388	-	7471B	PPB	08 Jul 2024 10:10:53	U	SMPL
P3076-06 - 1	7-W-3	7159	1.2678	-	7471B	PPB	08 Jul 2024 10:13:09	U	SMPL
P3076-08 - 1	7-E-2	7760	1.3730	-	7471B	PPB	08 Jul 2024 10:15:26	U	SMPL
P3076-09 - 1	7-E-3	6039	1.0718	-	7471B	PPB	08 Jul 2024 10:17:44	U	SMPL
P3076-11 - 1	7-S-2	5689	1.0105	-	7471B	PPB	08 Jul 2024 10:20:02	U	SMPL
P3076-12 - 1	7-S-3	6753	1.1968	-	7471B	PPB	08 Jul 2024 10:22:20	U	SMPL
P3076-14 - 1	7-B-2	6090	1.0807	-	7471B	PPB	08 Jul 2024 10:24:38	U	SMPL
P3076-15 - 1	7-B-3	5778	1.0261	-	7471B	PPB	08 Jul 2024 10:26:57	U	SMPL
PB161942BL - 1	PBS	-19	0.0116	-	7471B	PPB	08 Jul 2024 10:29:16	U	SMPL
PB161942BS - 1	LCSS	23098	4.0574	-	7471B	PPB	08 Jul 2024 10:31:34	U	SMPL
CCV48 - 1	CCV48	25882	4.5446	-	7471B	PPB	08 Jul 2024 10:33:50	U	SMPL
CCB48 - 1	CCB48	-158	-0.0128	-	7471B	PPB	08 Jul 2024 10:36:08	U	SMPL
P3077-02 - 1	1-N-2	7266	1.2865	-	7471B	PPB	08 Jul 2024 10:38:27	U	SMPL
P3077-03 - 1	1-N-3	7988	1.4129	-	7471B	PPB	08 Jul 2024 10:40:43	U	SMPL
P3077-05 - 1	1-W-2	5671	1.0074	-	7471B	PPB	08 Jul 2024 10:43:01	U	SMPL
P3077-06 - 1	1-W-2	5843	1.0375	-	7471B	PPB	08 Jul 2024 10:45:20	U	SMPL
P3077-08 - 1	1-S-2	6075	1.0781	-	7471B	PPB	08 Jul 2024 10:47:38	U	SMPL
P3077-09 - 1	1-S-3	9282	1.6359	-	7471B	PPB	08 Jul 2024 10:49:56	U	SMPL
P3077-11 - 1	1-B-2	4649	0.8285	-	7471B	PPB	08 Jul 2024 10:52:15	U	SMPL
P3077-12 - 1	1-B-3	6411	1.1369	-	7471B	PPB	08 Jul 2024 10:54:34	U	SMPL
P3157-01 - 1	PALISADES-POLE-YARD-DG-PAINT	484528	HIGH	-	7471B	PPB	08 Jul 2024 10:56:52	X	
P3157-04 - 1	PALISADES-POLE-YARD-LG-PAINT	1109831	HIGH	-	7471B	PPB	08 Jul 2024 10:58:42	X	
CCV49 - 1	CCV49	28888	5.0707	-	7471B	PPB	08 Jul 2024 11:04:50	U	SMPL
CCB49 - 1	CCB49	-408	-0.0565	-	7471B	PPB	08 Jul 2024 11:07:07	U	SMPL
P3157-07 - 1	PALISADES-POLE-YARD-Y-PAINT	878516	HIGH	-	7471B	PPB	08 Jul 2024 11:09:26	X	
P3157-10 - 1	PALISADES-POLE-YARD-W-PAINT	720118	HIGH	-	7471B	PPB	08 Jul 2024 11:11:13	X	
P3159-01 - 1	SU-04-070324	1637	0.3014	-	7471B	PPB	08 Jul 2024 11:19:51	U	SMPL
P3159-01DUP - 1	SU-04-070324DUP	1874	0.3429	-	7471B	PPB	08 Jul 2024 11:28:02	U	SMPL
P3159-01MS - 1	SU-04-070324MS	26909	4.7243	-	7471B	PPB	08 Jul 2024 11:30:18	U	SMPL
P3159-01MSD - 1	SU-04-070324MSD	27473	4.8231	-	7471B	PPB	08 Jul 2024 11:32:34	U	SMPL
P3166-01 - 1	NB-08-070324	3158	0.5676	-	7471B	PPB	08 Jul 2024 11:34:54	U	SMPL
P3157-01DLX10 - 1		72767	12.7501	-	7471B	PPB	08 Jul 2024 11:37:14	U	SMPL
P3157-04DLX10 - 1		307109	53.7634	-	7471B	PPB	08 Jul 2024 11:39:32	U	SMPL
P3157-07DLX10 - 1		183059	32.0528	-	7471B	PPB	08 Jul 2024 11:41:58	U	SMPL
CCV50 - 1	CCV50	29491	5.1762	-	7471B	PPB	08 Jul 2024 11:47:45	U	SMPL
CCB50 - 1	CCB50	-377	-0.0511	-	7471B	PPB	08 Jul 2024 11:50:02	U	SMPL
P3157-10DLX10 - 1		138665	24.2832	-	7471B	PPB	08 Jul 2024 11:53:53	U	SMPL
P3076-02A - 1		31103	5.4584	-	7471B	PPB	08 Jul 2024 11:58:38	U	SMPL
P3159-01LX5 - 1		-277	-0.0336	-	7471B	PPB	08 Jul 2024 12:00:55	U	SMPL
P3159-01A - 1		26879	4.7191	-	7471B	PPB	08 Jul 2024 12:03:16	U	SMPL
P3076-02LX5 - 1		1470	0.2722	-	7471B	PPB	08 Jul 2024 12:06:10	U	SMPL
P3157-01DL2X50 - 1		13073	2.3028	-	7471B	PPB	08 Jul 2024 12:08:48	U	SMPL
P3157-04DL2X100 - 1		31273	5.4881	-	7471B	PPB	08 Jul 2024 12:11:05	U	SMPL
P3157-07DL2X100 - 1		15559	2.7379	-	7471B	PPB	08 Jul 2024 12:13:24	U	SMPL
P3157-10DL2X100 - 1		12666	2.2316	-	7471B	PPB	08 Jul 2024 12:15:46	U	SMPL
CCV51 - 1	CCV51	28302	4.9681	-	7471B	PPB	08 Jul 2024 12:18:05	U	SMPL
CCB51 - 1	CCB51	-293	-0.0364	-	7471B	PPB	08 Jul 2024 12:20:24	U	SMPL