

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB131535

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	07/08/2024	14:50
2	CAL2	1	Water	NA	NA	20.2	7.01	07/08/2024	14:51
3	CAL3	1	Water	NA	NA	20.2	10.02	07/08/2024	14:53
4	ICV	1	Water	NA	NA	20.3	7.01	07/08/2024	14:55
5	CCV1	1	Water	NA	NA	20.2	2.01	07/08/2024	15:00
6	P3170-06	1	Water	NA	NA	21.2	9.22	07/08/2024	15:10
7	P3170-06DUP	1	Water	NA	NA	21.3	9.23	07/08/2024	15:11
8	CCV2	1	Water	NA	NA	20.3	12.03	07/08/2024	15:15

WORKLIST(Hardcopy Internal Chain)

UB 131535

WorkList Name : PH P3170

WorkList ID : 181688

Department : Wet-Chemistry

Date : 07-08-2024 14:35:14

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3170-06 B	WATER-DRUMS	Water	pH	Cool 4 deg C	PSEG03	D21	07/08/2024	9040C

Date/Time 07/08/24 14:45

Raw Sample Received by: JH WOC

Raw Sample Relinquished by: CP SN

Date/Time 07/08/24

Raw Sample Received by: CP SN

Raw Sample Relinquished by: JH WOC