


LB131538

SFAM01.1

INSTRUMENT ID: CV1

 μ Abs.:

37381



Linear

A= 0.0000e+000

B= 2.6341e-004 slope

C= 7.1543e-003 $y = mx + c$

Rho= 0.9995879

Accept = Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% of 1)
0.00	0.000	0.009	0.009	8	0.000	8					=
0.05	0.050					442	0				-2
0.20	0.200	0.197	-0.003	722	0.0 %	722					-4
2.50	2.500	2.401	-0.099	9088	0.0 %	9088					1
5.00	5.000	5.058	0.058	19177	0.0 %	19177					2
7.50	7.500	7.680	0.180	29131	0.0 %	29131					-1
10.0	10.000	9.854	-0.146	37381	0.0 %	37381					

LB131538

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	8	-		0 SFAM01.1	PPB	7/9/2024 11:38	Std
	0.2 S01	722	-		0.2 SFAM01.1	PPB	7/9/2024 11:40	Std
	2.5 S02	9088	-		2.5 SFAM01.1	PPB	7/9/2024 11:45	Std
	5 S03	19177	-		5 SFAM01.1	PPB	7/9/2024 11:47	Std
	7.5 S04	29131	-		7.5 SFAM01.1	PPB	7/9/2024 11:49	Std
	10 S05	37381	-		10 SFAM01.1	PPB	7/9/2024 11:52	Std
ICV021	ICV021	15008	3.9603 -		SFAM01.1	PPB	7/9/2024 11:55	SMPL
ICB021	ICB021	-252	-0.0592 -		SFAM01.1	PPB	7/9/2024 11:57	SMPL
CCV054	CCV054	18921	4.9911 -		SFAM01.1	PPB	7/9/2024 11:59	SMPL
CCB054	CCB054	-277	-0.0658 -		SFAM01.1	PPB	7/9/2024 12:01	SMPL
PB161971BL	PBW971	-6	0.0056 -		SFAM01.1	PPB	7/9/2024 12:04	SMPL
P2965-17	MA4D65	-19	0.0021 -		SFAM01.1	PPB	7/9/2024 12:06	SMPL
P3033-10	MA4D64	49	0.0201 -		SFAM01.1	PPB	7/9/2024 12:08	SMPL
CCV055	CCV055	18963	5.0021 -		SFAM01.1	PPB	7/9/2024 12:10	SMPL
CCB055	CCB055	-282	-0.0671 -		SFAM01.1	PPB	7/9/2024 12:13	SMPL