

LB131543


245.1

INSTRUMENT ID: CY1

Linear

μ Abs.:

29624



A= 0.0000e+000

B= 3.3448e-004

C= -3.5851e-002

Rho= 0.9997462

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	-0.027	-0.027	25	0.000	25					0/05
0.05	0.050	0.059	0.009	283	0.0 %	283	0				18
0.2	0.200	0.190	-0.010	675	0.0 %	675					5
2.5	2.500	2.499	-0.001	7579	0.0 %	7579					0
5.0	5.000	4.980	-0.020	14995	0.0 %	14995					0
7.5	7.500	7.677	0.177	23060	0.0 %	23060					2
10.0	10.000	9.873	-0.127	29624	0.0 %	29624					1

LB131543 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 09 Jul 2024 14:29:39

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	25	-	0.0000	245.1	PPB	09 Jul 2024 14:49:58	S	Std
0.05 - 1	50.05	283	-	0.0500	245.1	PPB	09 Jul 2024 14:52:16	S	Std
0.2 - 1	50.2	675	-	0.2000	245.1	PPB	09 Jul 2024 14:54:34	S	Std
2.5 - 1	52.5	7579	-	2.5000	245.1	PPB	09 Jul 2024 14:56:51	S	Std
5.0 - 1	55	14995	-	5.0000	245.1	PPB	09 Jul 2024 14:59:07	S	Std
7.5 - 1	57.5	23060	-	7.5000	245.1	PPB	09 Jul 2024 15:01:25	S	Std
10.0 - 1	510	29624	-	10.0000	245.1	PPB	09 Jul 2024 15:15:57	S	Std
ICV22 - 1	ICV22	11942	3.9585	-245.1		PPB	09 Jul 2024 15:23:59	U	SMPL
ICB22 - 1	ICB22	-32	-0.0466	-245.1		PPB	09 Jul 2024 15:28:48	U	SMPL
CCV56 - 1	CCV56	14117	4.6860	-245.1		PPB	09 Jul 2024 15:31:03	U	SMPL
CCB56 - 1	CCB56	-3	-0.0369	-245.1		PPB	09 Jul 2024 15:35:54	U	SMPL
CRA - 1	CRA	237	0.0434	-245.1		PPB	09 Jul 2024 15:38:09	U	SMPL
HighStd - 1	HighStd	29966	9.9871	-245.1		PPB	09 Jul 2024 15:40:23	U	SMPL
ChkStd - 1	ChkStd	18865	6.2741	-245.1		PPB	09 Jul 2024 15:42:39	U	SMPL
PB161972BS - 1	LCSW	10856	3.5952	-245.1		PPB	09 Jul 2024 15:47:16	U	SMPL
P3147-01 - 1	402	-176	-0.0947	-245.1		PPB	09 Jul 2024 15:49:33	U	SMPL
P3147-01DUP - 1	402DUP	-30	-0.0459	-245.1		PPB	09 Jul 2024 15:51:51	U	SMPL
P3147-01MS - 1	402MS	12251	4.0618	-245.1		PPB	09 Jul 2024 15:54:09	U	SMPL
P3147-01MSD - 1	402MSD	10302	3.4099	-245.1		PPB	09 Jul 2024 15:56:26	U	SMPL
CCV57 - 1	CCV57	14194	4.7117	-245.1		PPB	09 Jul 2024 16:04:15	U	SMPL
CCB57 - 1	CCB57	0	-0.0359	-245.1		PPB	09 Jul 2024 16:09:04	U	SMPL
P3147-01LX5 - 1		-60	-0.0559	-245.1		PPB	09 Jul 2024 16:11:19	U	SMPL
P3147-01A - 1		11027	3.6524	-245.1		PPB	09 Jul 2024 16:13:35	U	SMPL
PB161972BL - 1	PBW	-69	-0.0589	-245.1		PPB	09 Jul 2024 16:15:50	U	SMPL
CCV58 - 1	CCV58	15496	5.1472	-245.1		PPB	09 Jul 2024 16:18:07	U	SMPL
CCB58 - 1	CCB58	-22	-0.0432	-245.1		PPB	09 Jul 2024 16:22:57	U	SMPL