

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB131545

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3084

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.4	07/09/2024	08:45
2	P3170-06		1	12	100.00	>212.0	>212.0	07/09/2024	09:15
3	P3170-06DUP		1	12	100.00	>212.0	>212.0	07/09/2024	09:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

66131545

WorkList Name : fp-7-9

WorkList ID : 181723

Department : Wet-Chemistry

Date : 07-09-2024 08:07:20

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3170-06	WATER-DRUMS	Water	Flash Point	Cool 4 deg C	PSEG03	D21	07/08/2024	1010B

Date/Time 07/09/2024 08:36
 Raw Sample Received by: RM CWL
 Raw Sample Relinquished by: JH CWL

Date/Time 07/09/2024 10:30
 Raw Sample Received by: JH CWL
 Raw Sample Relinquished by: RM CWL