


LB131549

SFAM01.1

INSTRUMENT ID: CVZ

 μ Abs.:

47994



Linear

A= 0.0000e+000

B= 2.0777e-004 slope

C= 4.1022e-002 $r^2 = 0.9999$

Rho= 0.9999183

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	9.18
0.00	0.000	0.041	0.041	1	0.000	1					
0.05	0.050					442	0				
0.20	0.200	0.236	0.036	940	0.0 %	940					
2.50	2.500	2.406	-0.094	11382	0.0 %	11382					
5.00	5.000	4.977	-0.023	23755	0.0 %	23755					
7.50	7.500	7.527	0.027	36033	0.0 %	36033					
10.0	10.000	10.013	0.013	47994	0.0 %	47994					

9.18
= 18
- 4
0
0
0

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	1	-		0 SFAM01.1	PPB	7/10/2024 9:54	Std
	0.2 S01	940	-		0.2 SFAM01.1	PPB	7/10/2024 9:56	Std
	2.5 S02	11382	-		2.5 SFAM01.1	PPB	7/10/2024 9:58	Std
	5 S03	23755	-		5 SFAM01.1	PPB	7/10/2024 10:01	Std
	7.5 S04	36033	-		7.5 SFAM01.1	PPB	7/10/2024 10:03	Std
	10 S05	47994	-		10 SFAM01.1	PPB	7/10/2024 10:05	Std
ICV023	ICV023	18521	3.8891	-	SFAM01.1	PPB	7/10/2024 10:08	SMPL
ICB023	ICB023	-181	0.0034	-	SFAM01.1	PPB	7/10/2024 10:10	SMPL
CCV059	CCV059	24607	5.1535	-	SFAM01.1	PPB	7/10/2024 10:13	SMPL
CCB059	CCB059	-166	0.0065	-	SFAM01.1	PPB	7/10/2024 10:15	SMPL
PB161991BL	PBS991	40	0.0493	-	SFAM01.1	PPB	7/10/2024 10:17	SMPL
P2985-01	MA4CS4	24696	5.172	-	SFAM01.1	PPB	7/10/2024 10:19	SMPL
P2985-02	MA4CS4D	24679	5.1685	-	SFAM01.1	PPB	7/10/2024 10:22	SMPL
P2985-03	MA4CS4S	36007	7.5221	-	SFAM01.1	PPB	7/10/2024 10:24	SMPL
P2985-04	MA4CF9	474	0.1395	-	SFAM01.1	PPB	7/10/2024 10:26	SMPL
P2985-05	MA4CG0	242	0.0913	-	SFAM01.1	PPB	7/10/2024 10:29	SMPL
P2985-06	MA4D69	293172	HIGH	-	SFAM01.1	PPB	7/10/2024 10:31	
P2985-07	MA4BB8	4195	0.9126	-	SFAM01.1	PPB	7/10/2024 10:33	SMPL
P2985-08	MA4BB9	413	0.1268	-	SFAM01.1	PPB	7/10/2024 10:35	SMPL
P2985-09	MA4BC1	1973	0.4509	-	SFAM01.1	PPB	7/10/2024 10:37	SMPL
P2985-10	MA4BC2	171	0.0766	-	SFAM01.1	PPB	7/10/2024 10:40	SMPL
P2985-11	MA4BC7	1196	0.2895	-	SFAM01.1	PPB	7/10/2024 10:42	SMPL
P2985-12	MA4BC8	463	0.1372	-	SFAM01.1	PPB	7/10/2024 10:44	SMPL
P2985-13	MA4CW0	10113	2.1422	-	SFAM01.1	PPB	7/10/2024 10:46	SMPL
P2985-14	MA4CW1	4378	0.9506	-	SFAM01.1	PPB	7/10/2024 10:49	SMPL
P2985-15	MA4BC5	374	0.1187	-	SFAM01.1	PPB	7/10/2024 10:51	SMPL
P2985-16	MA4BC4	2430	0.5459	-	SFAM01.1	PPB	7/10/2024 10:53	SMPL
P2985-17	MA4CT8	3951	0.8619	-	SFAM01.1	PPB	7/10/2024 10:55	SMPL
P2985-18	MA4CT9	1666	0.3872	-	SFAM01.1	PPB	7/10/2024 10:58	SMPL
P2985-19	MA4CW3	887	0.2253	-	SFAM01.1	PPB	7/10/2024 11:00	SMPL
CCV060	CCV060	24062	5.0403	-	SFAM01.1	PPB	7/10/2024 11:02	SMPL
CCB060	CCB060	-120	0.0161	-	SFAM01.1	PPB	7/10/2024 11:05	SMPL
P2985-20	MA4CW4	120	0.066	-	SFAM01.1	PPB	7/10/2024 11:07	SMPL
P2985-21	MA4CW6	4132	0.8995	-	SFAM01.1	PPB	7/10/2024 11:09	SMPL
P2985-22	MA4B68	2279	0.5145	-	SFAM01.1	PPB	7/10/2024 11:11	SMPL
P2985-23	MA4B69	2736	0.6095	-	SFAM01.1	PPB	7/10/2024 11:14	SMPL
PB161992BL	PBS992	-29	0.035	-	SFAM01.1	PPB	7/10/2024 11:16	SMPL
P2986-01	MA4CT2	7575	1.6149	-	SFAM01.1	PPB	7/10/2024 11:18	SMPL
P2986-02	MA4CT2D	7653	1.6311	-	SFAM01.1	PPB	7/10/2024 11:20	SMPL
P2986-03	MA4CT2S	18254	3.8336	-	SFAM01.1	PPB	7/10/2024 11:23	SMPL
P2986-04	MA4B98	2817	0.6263	-	SFAM01.1	PPB	7/10/2024 11:25	SMPL
P2986-05	MA4BB3	1016	0.2521	-	SFAM01.1	PPB	7/10/2024 11:27	SMPL
P2986-06	MA4BB4	149	0.072	-	SFAM01.1	PPB	7/10/2024 11:30	SMPL
P2986-07	MA4BB5	180	0.0784	-	SFAM01.1	PPB	7/10/2024 11:32	SMPL
P2986-08	MA4CG4	164	0.0751	-	SFAM01.1	PPB	7/10/2024 11:34	SMPL
P2986-09	MA4CG5	78	0.0572	-	SFAM01.1	PPB	7/10/2024 11:36	SMPL

INSTRUMENT ID : CV1

P2986-10	MA4BD0	6997	1.4948 -	SFAM01.1	PPB	7/10/2024 11:39	SMPL
P2986-11	MA4CW5	13054	2.7532 -	SFAM01.1	PPB	7/10/2024 11:41	SMPL
P2986-12	MA4CW9	9872	2.0921 -	SFAM01.1	PPB	7/10/2024 11:43	SMPL
P2986-13	MA4CX0	5572	1.1987 -	SFAM01.1	PPB	7/10/2024 11:46	SMPL
P2986-14	MA4CX1	2040	0.4649 -	SFAM01.1	PPB	7/10/2024 11:48	SMPL
P2986-15	MA4CX3	181	0.0786 -	SFAM01.1	PPB	7/10/2024 11:50	SMPL
CCV061	CCV061	24122	5.0528 -	SFAM01.1	PPB	7/10/2024 11:52	SMPL
CCB061	CCB061	-130	0.014 -	SFAM01.1	PPB	7/10/2024 11:55	SMPL
P2986-16	MA4CX4	1241	0.2989 -	SFAM01.1	PPB	7/10/2024 11:57	SMPL
P2986-17	MA4CY1	6522	1.3961 -	SFAM01.1	PPB	7/10/2024 11:59	SMPL
P2986-18	MA4B65	6461	1.3834 -	SFAM01.1	PPB	7/10/2024 12:01	SMPL
P2986-19	MA4B66	1059	0.261 -	SFAM01.1	PPB	7/10/2024 12:04	SMPL
P2986-20	MA4B67	178	0.078 -	SFAM01.1	PPB	7/10/2024 12:06	SMPL
P2986-21	MA4CX5	660	0.1781 -	SFAM01.1	PPB	7/10/2024 12:08	SMPL
P2986-22	MA4CX7	11212	2.3705 -	SFAM01.1	PPB	7/10/2024 12:11	SMPL
P2986-23	MA4D71	387414	HIGH -	SFAM01.1	PPB	7/10/2024 12:13	
CCV062	CCV062	23702	4.9655 -	SFAM01.1	PPB	7/10/2024 12:26	SMPL
CCB062	CCB062	146	0.0714 -	SFAM01.1	PPB	7/10/2024 12:29	SMPL
P2985-06DLX25	MA4D69	21176	4.4407 -	SFAM01.1	PPB	7/10/2024 12:31	SMPL
P2986-23DLX25	MA4D71	25095	5.2549 -	SFAM01.1	PPB	7/10/2024 12:40	SMPL
CCV063	CCV063	22275	4.669 -	SFAM01.1	PPB	7/10/2024 12:51	SMPL
CCB063	CCB063	-56	0.0294 -	SFAM01.1	PPB	7/10/2024 12:53	SMPL