

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB131572

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	07/11/2024	10:50
2	CAL2	1	Water	NA	NA	20.3	7.00	07/11/2024	10:41
3	CAL3	1	Water	NA	NA	20.3	10.02	07/11/2024	10:44
4	ICV	1	Water	NA	NA	20.3	7.02	07/11/2024	10:45
5	CCV1	1	Water	NA	NA	20.3	2.01	07/11/2024	10:47
6	P3208-12	1	Water	NA	NA	23.6	7.52	07/11/2024	10:50
7	P3208-12DUP	1	Water	NA	NA	23.6	7.53	07/11/2024	10:51
8	P3210-03	1	Water	NA	NA	23.0	8.81	07/11/2024	11:00
9	P3210-04	1	Water	NA	NA	22.5	6.12	07/11/2024	11:10
10	P3210-05	1	Water	NA	NA	22.8	6.73	07/11/2024	11:15
11	P3212-02	1	Water	NA	NA	20.4	9.03	07/11/2024	11:22
12	CCV2	1	Water	NA	NA	20.3	12.02	07/11/2024	11:25

WORKLIST(Hardcopy Internal Chain)

UP 131572

WorkList Name : PH P3212 WATER

WorkList ID : 181774

Department : Wet-Chemistry

Date : 07-11-2024 10:38:05

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3208-12	A 3078	Water	pH	Cool 4 deg C	PSEG03	C21	07/10/2024	9040C
P3210-03	A D255	Water	pH	Cool 4 deg C	PSEG03	C21	07/10/2024	9040C
P3210-04	A D256	Water	pH	Cool 4 deg C	PSEG03	C21	07/10/2024	9040C
P3210-05	A D257	Water	pH	Cool 4 deg C	PSEG03	C21	07/10/2024	9040C
P3212-02	A 407	Water	pH	Cool 4 deg C	PSEG03	C31	07/10/2024	9040C

Date/Time 07-11-24 10:40

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 07-11-24 12:40

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]