


LB131574

7471B

INSTRUMENT ID: CV1

μ Abs.:

57561



Linear

A= 0.0000e+000

B= 1.7253e-004

C= -4.7083e-002

Rho= 0.9997665

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	9/6/1)
0.0	0.000	-0.033	-0.033	83	0.000	83					
0.2	0.200	0.210	0.010	1490	0.0 %	1490					5
2.5	2.500	2.441	-0.059	14420	0.0 %	14420					-2
5.0	5.000	5.087	0.087	29756	0.0 %	29756					2
7.5	7.500	7.612	0.112	44391	0.0 %	44391					1
10.0	10.000	9.884	-0.116	57561	0.0 %	57561					-)

LB131574 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 11 Jul 2024 12:47:17

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	83	-	0.0000	7471B	PPB	11 Jul 2024 13:42:31	S	Std
0.2 - 1	50.2	1490	-	0.2000	7471B	PPB	11 Jul 2024 13:44:48	S	Std
2.5 - 1	52.5	14420	-	2.5000	7471B	PPB	11 Jul 2024 13:47:06	S	Std
5.0 - 1	55	29756	-	5.0000	7471B	PPB	11 Jul 2024 13:49:23	S	Std
7.5 - 1	57.5	44391	-	7.5000	7471B	PPB	11 Jul 2024 13:51:42	S	Std
10.0 - 1	50	57561	-	10.0000	7471B	PPB	11 Jul 2024 13:54:01	S	Std
ICV25 - 1	ICV25	23721	4.0454	-	7471B	PPB	11 Jul 2024 13:57:02	U	SMPL
ICB25 - 1	ICB25	101	-0.0297	-	7471B	PPB	11 Jul 2024 14:01:58	U	SMPL
CCV67 - 1	CCV67	29931	5.1168	-	7471B	PPB	11 Jul 2024 14:04:13	U	SMPL
CCB67 - 1	CCB67	145	-0.0221	-	7471B	PPB	11 Jul 2024 14:09:07	U	SMPL
CRA - 1	CRA	1459	0.2046	-	7471B	PPB	11 Jul 2024 14:11:22	U	SMPL
HighStd - 1	HighStd	57450	9.8646	-	7471B	PPB	11 Jul 2024 14:13:37	U	SMPL
ChkStd - 1	ChkStd	40129	6.8763	-	7471B	PPB	11 Jul 2024 14:18:42	U	SMPL
PB162014BL - 1	PBS	270	-0.0005	-	7471B	PPB	11 Jul 2024 14:20:57	U	SMPL
PB162014BS - 1	LCSS	23961	4.0869	-	7471B	PPB	11 Jul 2024 14:23:26	U	SMPL
P3170-01 - 1	SOIL-DRUMS-1	3854	0.6178	-	7471B	PPB	11 Jul 2024 14:25:42	U	SMPL
P3170-03 - 1	SOIL-DRUMS-2	3158	0.4978	-	7471B	PPB	11 Jul 2024 14:28:01	U	SMPL
P3181-01 - 1	OR-02-070924	2043	0.3054	-	7471B	PPB	11 Jul 2024 14:30:22	U	SMPL
P3182-01 - 1	SG-1	764	0.0847	-	7471B	PPB	11 Jul 2024 14:32:41	U	SMPL
P3183-01 - 1	N28660	8826	1.4756	-	7471B	PPB	11 Jul 2024 14:34:57	U	SMPL
CCV68 - 1	CCV68	28507	4.8712	-	7471B	PPB	11 Jul 2024 14:37:13	U	SMPL
CCB68 - 1	CCB68	-188	-0.0795	-	7471B	PPB	11 Jul 2024 14:39:31	U	SMPL
P3183-05 - 1	VNJ-226	3554	0.5661	-	7471B	PPB	11 Jul 2024 14:41:49	U	SMPL
P3183-07 - 1	VNJ-229	2213	0.3347	-	7471B	PPB	11 Jul 2024 14:44:05	U	SMPL
P3184-01 - 1	COMP(00187-00189-00186)	1799	0.2633	-	7471B	PPB	11 Jul 2024 14:46:22	U	SMPL
P3184-04 - 1	MOO-23-00199	2149	0.3237	-	7471B	PPB	11 Jul 2024 14:48:38	U	SMPL
P3186-01 - 1	WC-4	5156	0.8425	-	7471B	PPB	11 Jul 2024 14:50:55	U	SMPL
P3187-01 - 1	WC-1	3977	0.6391	-	7471B	PPB	11 Jul 2024 14:53:12	U	SMPL
P3187-05 - 1	WC-2	2551	0.3930	-	7471B	PPB	11 Jul 2024 14:55:30	U	SMPL
P3187-09 - 1	WC-3	5861	0.9641	-	7471B	PPB	11 Jul 2024 14:57:47	U	SMPL
P3193-02 - 1	FILL-MATERIAL-1	5185	0.8475	-	7471B	PPB	11 Jul 2024 15:00:04	U	SMPL
P3193-04 - 1	FILL-MATERIAL-2	2084	0.3125	-	7471B	PPB	11 Jul 2024 15:02:22	U	SMPL
CCV69 - 1	CCV69	29064	4.9673	-	7471B	PPB	11 Jul 2024 15:04:38	U	SMPL
CCB69 - 1	CCB69	-171	-0.0766	-	7471B	PPB	11 Jul 2024 15:07:02	U	SMPL
P3193-06 - 1	FILL-MATERIAL-3	5620	0.9225	-	7471B	PPB	11 Jul 2024 15:09:22	U	SMPL
P3193-06DUP - 1	FILL-MATERIAL-3DUP	5064	0.8266	-	7471B	PPB	11 Jul 2024 15:11:37	U	SMPL
P3193-06MS - 1	FILL-MATERIAL-3MS	29254	5.0000	-	7471B	PPB	11 Jul 2024 15:13:56	U	SMPL
P3193-06MSD - 1	FILL-MATERIAL-3MSD	29726	5.0815	-	7471B	PPB	11 Jul 2024 15:16:12	U	SMPL
P3208-01 - 1	352	2846	0.4439	-	7471B	PPB	11 Jul 2024 15:18:31	U	SMPL
P3208-03 - 1	VNJ-207	2800	0.4360	-	7471B	PPB	11 Jul 2024 15:20:53	U	SMPL
P3208-05 - 1	CONC-PADS	1027	0.1301	-	7471B	PPB	11 Jul 2024 15:23:11	U	SMPL
P3193-06LX5 - 1		1022	0.1292	-	7471B	PPB	11 Jul 2024 15:25:29	U	SMPL
P3193-06A - 1		29545	5.0502	-	7471B	PPB	11 Jul 2024 15:27:46	U	SMPL
CCV70 - 1	CCV70	29325	5.0123	-	7471B	PPB	11 Jul 2024 15:30:01	U	SMPL
CCB70 - 1	CCB70	31	-0.0417	-	7471B	PPB	11 Jul 2024 15:48:58	U	SMPL