

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB131578

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	07/11/2024	11:00
2	CAL2	1	Water	NA	NA	20.2	7.00	07/11/2024	11:01
3	CAL3	1	Water	NA	NA	20.2	10.02	07/11/2024	11:05
4	ICV	1	Water	NA	NA	20.3	7.02	07/11/2024	11:10
5	CCV1	1	Water	NA	NA	20.3	2.01	07/11/2024	11:15
6	P3195-01	1	Solid	20.02	20	22.8	6.91	07/11/2024	11:20
7	P3223-01	1	Solid	20.03	20	24.3	9.64	07/11/2024	11:30
8	P3223-01DUP	1	Solid	20.04	20	24.4	9.65	07/11/2024	11:31
9	CCV2	1	Water	NA	NA	20.2	12.02	07/11/2024	11:35

UB 131578

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH P3212 WorkList ID : 181773 Department : Wet-Chemistry Date : 07-11-2024 10:37:05

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3195-01	34806	Solid	pH	Cool 4 deg C	PSEG03	C21	07/10/2024	9045D
P3223-01	PILE-CAP	Solid	pH	Cool 4 deg C	TULL02	C31	07/11/2024	9045D

Date/Time07-11-2410:55

Raw Sample Received by:ZO WUCY

Raw Sample Relinquished by:ZO WUCY

Date/Time07-11-2413:00

Raw Sample Received by:ZO WUCY

Raw Sample Relinquished by:ZO WUCY