

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB131584

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3084

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.1	07/11/2024	10:40
2	P3206-01		1	13	100.00	>212.0	>212.0	07/11/2024	11:10
3	P3206-01DUP		1	13	100.00	>212.0	>212.0	07/11/2024	11:40
4	P3206-02		1	14	100.00	>212.0	>212.0	07/11/2024	12:10
5	P3208-12		1	11	100.00	>212.0	>212.0	07/11/2024	12:40
6	P3210-03		1	14	100.00	>212.0	>212.0	07/11/2024	13:10
7	P3210-04		1	11	100.00	>212.0	>212.0	07/11/2024	13:40
8	P3210-05		1	14	100.00	>212.0	>212.0	07/11/2024	14:10
9	P3212-02		1	12	100.00	>212.0	>212.0	07/11/2024	14:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6131584

WorkList Name : fp-7-11 WorkList ID : 181760 Department : Wet-Chemistry Date : 07-11-2024 07:48:09

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3206-01 B	626-B	Water	Flash Point	Cool 4 deg C	PSEG03	C31	07/10/2024	1010B
P3206-02 B	626-A	Water	Flash Point	Cool 4 deg C	PSEG03	C31	07/10/2024	1010B
P3208-12 B	3078	Water	Flash Point	Cool 4 deg C	PSEG03	C21	07/10/2024	1010B
P3210-03 B	D255	Water	Flash Point	Cool 4 deg C	PSEG03	C21	07/10/2024	1010B
P3210-04 B	D256	Water	Flash Point	Cool 4 deg C	PSEG03	C21	07/10/2024	1010B
P3210-05 B	D257	Water	Flash Point	Cool 4 deg C	PSEG03	C21	07/10/2024	1010B
P3212-02 B	407	Water	Flash Point	Cool 4 deg C	PSEG03	C31	07/10/2024	1010B

Date/Time 07/11/2024 10:30
Raw Sample Received by: RM CWL
Raw Sample Relinquished by: ALOCC

Date/Time 07/11/2024 15:15
Raw Sample Received by: ALOCC
Raw Sample Relinquished by: RM CWL