

LB191599


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

47818



A= 0.0000e+000

B= 2.0991e-004 *slope*C= -3.0103e-003 *y-intercept*

Rho= 0.9999783

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.012	0.012	72	0.000	72					
0.05	0.050					442	0				
0.20	0.200	0.195	-0.005	944	0.0 %	944					
2.50	2.500	2.517	0.017	12006	0.0 %	12006					
5.00	5.000	4.979	-0.021	23736	0.0 %	23736					
7.50	7.500	7.462	-0.038	35561	0.0 %	35561					
10.0	10.000	10.035	0.035	47818	0.0 %	47818					

11
13
10
0

LB131599

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	72	-		0 SFAM01.1	PPB	7/12/2024 14:18	Std
	0.2 S01	944	-		0.2 SFAM01.1	PPB	7/12/2024 14:20	Std
	2.5 S02	12006	-		2.5 SFAM01.1	PPB	7/12/2024 14:22	Std
	5 S03	23736	-		5 SFAM01.1	PPB	7/12/2024 14:24	Std
	7.5 S04	35561	-		7.5 SFAM01.1	PPB	7/12/2024 14:27	Std
	10 S05	47818	-		10 SFAM01.1	PPB	7/12/2024 14:29	Std
ICV27	ICV027	19557	4.1022	-	SFAM01.1	PPB	7/12/2024 14:32	SMPL
ICB027	ICB027	-66	-0.0169	-	SFAM01.1	PPB	7/12/2024 14:34	SMPL
CCV073	CCV073	24179	5.0724	-	SFAM01.1	PPB	7/12/2024 14:37	SMPL
CCB073	CCB073	-92	-0.0223	-	SFAM01.1	PPB	7/12/2024 14:39	SMPL
PB162045BL	PBS045	35	0.0043	-	SFAM01.1	PPB	7/12/2024 14:41	SMPL
P3033-01	MA4CF0	9429	1.9762	-	SFAM01.1	PPB	7/12/2024 14:43	SMPL
P3033-02	MA4CG2	30913	6.486	-	SFAM01.1	PPB	7/12/2024 14:51	SMPL
P3033-03	MA4CG2D	29876	6.2683	-	SFAM01.1	PPB	7/12/2024 14:53	SMPL
P3033-05	MA4CG3	14475	3.0354	-	SFAM01.1	PPB	7/12/2024 14:58	SMPL
P3033-06	MA4B96	3863	0.8079	-	SFAM01.1	PPB	7/12/2024 15:00	SMPL
P3033-07	MA4B97	11385	2.3868	-	SFAM01.1	PPB	7/12/2024 15:02	SMPL
P3033-08	MA4CB8	6060	1.269	-	SFAM01.1	PPB	7/12/2024 15:05	SMPL
P3033-09	MA4CB9	1459	0.3032	-	SFAM01.1	PPB	7/12/2024 15:07	SMPL
P3033-11	MA4D72	377995	HIGH	-	SFAM01.1	PPB	7/12/2024 15:09	
P3033-12	MA4CC0	319	0.064	-	SFAM01.1	PPB	7/12/2024 15:11	SMPL
P3033-13	MA4CC4	1424	0.2959	-	SFAM01.1	PPB	7/12/2024 15:13	SMPL
P3033-14	MA4CC9	1126	0.2333	-	SFAM01.1	PPB	7/12/2024 15:16	SMPL
P3033-15	MA4CD0	193	0.0375	-	SFAM01.1	PPB	7/12/2024 15:18	SMPL
PB162046BL	PBS046	34	0.0041	-	SFAM01.1	PPB	7/12/2024 15:20	SMPL
P3036-01	MA4CW2	10229	2.1442	-	SFAM01.1	PPB	7/12/2024 15:22	SMPL
P3036-02	MA4CW2D	8624	1.8073	-	SFAM01.1	PPB	7/12/2024 15:25	SMPL
P3036-03	MA4CW2S	21189	4.4448	-	SFAM01.1	PPB	7/12/2024 15:27	SMPL
P3036-04	MA4CD2	45	0.0064	-	SFAM01.1	PPB	7/12/2024 15:29	SMPL
CCV074	CCV074	23896	5.013	-	SFAM01.1	PPB	7/12/2024 15:31	SMPL
CCB074	CCB074	-176	-0.04	-	SFAM01.1	PPB	7/12/2024 15:34	SMPL
P3036-05	MA4CD3	5128	1.0734	-	SFAM01.1	PPB	7/12/2024 15:36	SMPL
P3036-06	MA4CD4	3125	0.653	-	SFAM01.1	PPB	7/12/2024 15:38	SMPL
P3036-07	MA4CD7	2496	0.5209	-	SFAM01.1	PPB	7/12/2024 15:41	SMPL
P3036-08	MA4D73	260732	54.7275	-	SFAM01.1	PPB	7/12/2024 15:43	SMPL
P3036-09	MA4CE2	2689	0.5614	-	SFAM01.1	PPB	7/12/2024 15:45	SMPL
P3036-10	MA4CE4	1392	0.2892	-	SFAM01.1	PPB	7/12/2024 15:48	SMPL
P3036-11	MA4CE8	6574	1.3769	-	SFAM01.1	PPB	7/12/2024 15:50	SMPL
P3033-04	MA4CG2S	41616	8.7326	-	SFAM01.1	PPB	7/12/2024 15:52	SMPL
CCV075	CCV075	22646	4.7506	-	SFAM01.1	PPB	7/12/2024 16:15	SMPL
CCB075	CCB075	-136	-0.0316	-	SFAM01.1	PPB	7/12/2024 16:18	SMPL
P3033-11DLX25	MA4D72	24827	5.2084	-	SFAM01.1	PPB	7/12/2024 16:25	SMPL
P3036-08DLX10	MA4D73	29877	6.2685	-	SFAM01.1	PPB	7/12/2024 16:32	SMPL
CCV076	CCV076	22401	4.6992	-	SFAM01.1	PPB	7/12/2024 16:37	SMPL
CCB076	CCB076	14	-0.0001	-	SFAM01.1	PPB	7/12/2024 16:39	SMPL