

Analytical Summary Report

Analysis Method: 9040C

Analyst By : Rubina

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB131600

Slope : 98.6

pH Meter ID : WC pH meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	07/12/2024	15:00
2	CAL2	1	Water	NA	NA	20.4	6.99	07/12/2024	15:02
3	CAL3	1	Water	NA	NA	20.4	10.02	07/12/2024	15:05
4	ICV	1	Water	NA	NA	20.3	7.01	07/12/2024	15:08
5	CCV1	1	Water	NA	NA	20.4	2.01	07/12/2024	15:13
6	P3236-01	1	Water	NA	NA	24.5	4.97	07/12/2024	15:15
7	P3236-01DUP	1	Water	NA	NA	24.4	4.96	07/12/2024	15:18
8	CCV2	1	Water	NA	NA	20.4	11.8	07/12/2024	15:22

WORKLIST(Hardcopy Internal Chain)

LB 13600

WorkList Name : PH-071224

WorkList ID : 181824

Department : Wet-Chemistry

Date : 07-12-2024 14:50:53

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3236-01	JS9600	Water	pH	Cool 4 deg C	STER17	K11	06/24/2024	9040C

Date/Time 07/12/24 16:55
Raw Sample Received by: RM CWC
Raw Sample Relinquished by: CP Jm

Date/Time 07/12/24 15:30
Raw Sample Received by: CP Jm
Raw Sample Relinquished by: RM CWC